

LB125665


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4340E



A= 0.0000e+000

B= 2.3078e-004

C= -1.1508e-001

Rho= 0.9992514

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	711	0.000	711	0				13
0.2	0.200	0.226	0.026	1477	0.0 %	1477					7/0
2.5	2.500	2.256	-0.244	10272	0.0 %	10272					4
5.0	5.000	5.218	0.218	23107	0.0 %	23107					1
7.5	7.500	7.550	0.050	33213	0.0 %	33213					1
10.0	10.000	9.902	-0.098	43405	0.0 %	43405					1

LB125665

Method: 7470A

Operator: Admin

Date of Analysis: 26 May 2023 11:09:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	711	-	0.0000	7470A	PPB	26 May 2023 11:22:21	S	Std
0.2 - 1	50.2	1477	-	0.2000	7470A	PPB	26 May 2023 11:24:39	S	Std
2.5 - 1	52.5	10272	-	2.5000	7470A	PPB	26 May 2023 11:26:56	S	Std
5.0 - 1	55	23107	-	5.0000	7470A	PPB	26 May 2023 11:34:55	S	Std
7.5 - 1	57.5	33213	-	7.5000	7470A	PPB	26 May 2023 11:39:49	S	Std
10.0 - 1	510	43405	-	10.0000	7470A	PPB	26 May 2023 11:44:47	S	Std
ICV94 - 1	ICV94	18474	4.1484	-	7470A	PPB	26 May 2023 11:47:34	U	SMPL
ICB94 - 1	ICB94	478	-0.0048	-	7470A	PPB	26 May 2023 11:49:49	U	SMPL
CCV26 - 1	CCV26	23675	5.3487	-	7470A	PPB	26 May 2023 11:52:06	U	SMPL
CCB26 - 1	CCB26	493	-0.0013	-	7470A	PPB	26 May 2023 11:54:21	U	SMPL
CRA - 1	CRA	1489	0.2286	-	7470A	PPB	26 May 2023 11:56:38	U	SMPL
HighStd - 1	HighStd	42524	9.6988	-	7470A	PPB	26 May 2023 11:58:53	U	SMPL
ChkStd - 1	ChkStd	30898	7.0157	-	7470A	PPB	26 May 2023 12:01:09	U	SMPL
PB153058BL - 1	PBW	473	-0.0059	-	7470A	PPB	26 May 2023 12:03:27	U	SMPL
PB153058BS - 1	LCSW	17694	3.9684	-	7470A	PPB	26 May 2023 12:05:45	U	SMPL
O2899-03 - 1	WC06	576	0.0178	-	7470A	PPB	26 May 2023 12:08:01	U	SMPL
O2899-05 - 1	WC01	617	0.0273	-	7470A	PPB	26 May 2023 12:10:19	U	SMPL
O2899-07 - 1	WC08	704	0.0474	-	7470A	PPB	26 May 2023 12:12:36	U	SMPL
O2899-09 - 1	WC14	681	0.0421	-	7470A	PPB	26 May 2023 12:14:54	U	SMPL
O2908-03 - 1	WC10	697	0.0458	-	7470A	PPB	26 May 2023 12:17:09	U	SMPL
CCV27 - 1	CCV27	23153	5.2283	-	7470A	PPB	26 May 2023 12:19:25	U	SMPL
CCB27 - 1	CCB27	532	0.0077	-	7470A	PPB	26 May 2023 12:21:40	U	SMPL
O2908-05 - 1	WC09	675	0.0407	-	7470A	PPB	26 May 2023 12:23:57	U	SMPL
O2908-08 - 1	WC13	656	0.0363	-	7470A	PPB	26 May 2023 12:26:13	U	SMPL
O2913-02 - 1	TP-70A	784	0.0659	-	7470A	PPB	26 May 2023 12:28:28	U	SMPL
O2916-02 - 1	SU-621-COMP-01	1056	0.1286	-	7470A	PPB	26 May 2023 12:30:44	U	SMPL
O2919-03 - 1	IMPACTED-STONE	738	0.0552	-	7470A	PPB	26 May 2023 12:33:00	U	SMPL
O2919-03DUP - 1	IMPACTED-STONEDUP	731	0.0536	-	7470A	PPB	26 May 2023 12:35:16	U	SMPL
O2919-03MS - 1	IMPACTED-STONEMS	20149	4.5350	-	7470A	PPB	26 May 2023 12:37:33	U	SMPL
O2919-03MSD - 1	IMPACTED-STONEMSD	19475	4.3794	-	7470A	PPB	26 May 2023 12:39:50	U	SMPL
PB153061BL - 1	PBW	515	0.0038	-	7470A	PPB	26 May 2023 12:42:08	U	SMPL
PB153061BS - 1	LCSW	18545	4.1648	-	7470A	PPB	26 May 2023 12:44:25	U	SMPL
CCV28 - 1	CCV28	23867	5.3930	-	7470A	PPB	26 May 2023 12:46:41	U	SMPL
CCB28 - 1	CCB28	518	0.0045	-	7470A	PPB	26 May 2023 12:48:58	U	SMPL
O2899-12 - 1	TWP01	677	0.0412	-	7470A	PPB	26 May 2023 12:51:15	U	SMPL
O2908-09 - 1	TWP02	1016	0.1194	-	7470A	PPB	26 May 2023 12:58:27	U	SMPL
O2908-09DUP - 1	TWP02DUP	888	0.0899	-	7470A	PPB	26 May 2023 13:00:43	U	SMPL
O2908-09MS - 1	TWP02MS	19108	4.2947	-	7470A	PPB	26 May 2023 13:02:58	U	SMPL
O2908-09MSD - 1	TWP02MSD	18165	4.0771	-	7470A	PPB	26 May 2023 13:05:14	U	SMPL
O2908-12 - 1	TWP02	718	0.0506	-	7470A	PPB	26 May 2023 13:07:32	U	SMPL
O2945-08 - 1	TWP03	1538	0.2399	-	7470A	PPB	26 May 2023 13:09:50	U	SMPL
O2945-10 - 1	TWP04	824	0.0751	-	7470A	PPB	26 May 2023 13:12:07	U	SMPL
O2945-13 - 1	TWP03	110	-0.0897	-	7470A	PPB	26 May 2023 13:14:24	U	SMPL
O2945-14 - 1	TWP04	174	-0.0749	-	7470A	PPB	26 May 2023 13:16:44	U	SMPL
CCV29 - 1	CCV29	22550	5.0891	-	7470A	PPB	26 May 2023 13:19:00	U	SMPL
CCB29 - 1	CCB29	534	0.0082	-	7470A	PPB	26 May 2023 13:21:15	U	SMPL
PB153024TB - 1	PB153024TB	678	0.0414	-	7470A	PPB	26 May 2023 13:23:32	U	SMPL
O2919-03LX5 - 1		178	-0.0740	-	7470A	PPB	26 May 2023 13:25:47	U	SMPL
O2919-03A - 1		17748	3.9809	-	7470A	PPB	26 May 2023 13:28:03	U	SMPL
O2908-09LX5 - 1		80	-0.0966	-	7470A	PPB	26 May 2023 13:30:19	U	SMPL
O2908-09A - 1		18970	4.2629	-	7470A	PPB	26 May 2023 13:32:36	U	SMPL
CCV30 - 1	CCV30	23664	5.3462	-	7470A	PPB	26 May 2023 13:34:53	U	SMPL
CCB30 - 1	CCB30	527	0.0065	-	7470A	PPB	26 May 2023 13:37:11	U	SMPL