


LB125674

7471B

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 1.9156e-004

C= -1.8659e-002

Rho= 0.9997324

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.015	0.015	176	0.000	176					6
0.2	0.200	0.212	0.012	1206	0.0 %	1206					-5
2.5	2.500	2.371	-0.129	12477	0.0 %	12477					2
5.0	5.000	5.103	0.103	26735	0.0 %	26735					51
7.5	7.500	7.590	0.090	39718	0.0 %	39718					
10.0	10.000	9.909	-0.091	51827	0.0 %	51827					

LB125674 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 May 2023 13:59:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	176	-	0.0000	7471B	PPB	26 May 2023 14:01:55	S	Std
0.2 - 1	50.2	1206	-	0.2000	7471B	PPB	26 May 2023 14:04:13	S	Std
2.5 - 1	52.5	12477	-	2.5000	7471B	PPB	26 May 2023 14:06:30	S	Std
5.0 - 1	53	26735	-	5.0000	7471B	PPB	26 May 2023 14:13:56	S	Std
7.5 - 1	57.5	39718	-	7.5000	7471B	PPB	26 May 2023 14:18:49	S	Std
10.0 - 1	59	51827	-	10.0000	7471B	PPB	26 May 2023 14:26:15	S	Std
ICV95 - 1	ICV95	21147	4.0322	-	7471B	PPB	26 May 2023 14:31:20	U	SMPL
ICB95 - 1	ICB95	-106	-0.0390	-	7471B	PPB	26 May 2023 14:33:37	U	SMPL
CCV31 - 1	CCV31	27000	5.1533	-	7471B	PPB	26 May 2023 14:35:54	U	SMPL
CCB31 - 1	CCB31	-111	-0.0399	-	7471B	PPB	26 May 2023 14:38:09	U	SMPL
CRA - 1	CRA	1176	0.2066	-	7471B	PPB	26 May 2023 14:40:27	U	SMPL
HighStd - 1	HighStd	51787	9.9014	-	7471B	PPB	26 May 2023 14:42:43	U	SMPL
ChkStd - 1	ChkStd	36754	7.0218	-	7471B	PPB	26 May 2023 14:44:58	U	SMPL
PB153088BL - 1	PBS	-199	-0.0568	-	7471B	PPB	26 May 2023 14:47:17	U	SMPL
PB153088BS - 1	LCSS	23913	4.5620	-	7471B	PPB	26 May 2023 14:52:33	U	SMPL
O2916-01 - 1	SU-621-COMP-01	14243	2.7097	-	7471B	PPB	26 May 2023 14:54:49	U	SMPL
O2916-01DUP - 1	SU-621-COMP-01DUP	11830	2.2474	-	7471B	PPB	26 May 2023 14:57:06	U	SMPL
O2916-01MS - 1	SU-621-COMP-01MS	36447	6.9630	-	7471B	PPB	26 May 2023 14:59:25	U	SMPL
O2916-01MSD - 1	SU-621-COMP-01MSD	34835	6.6542	-	7471B	PPB	26 May 2023 15:01:44	U	SMPL
O2919-01 - 1	IMPACTED-STONE	1212	0.2135	-	7471B	PPB	26 May 2023 15:04:02	U	SMPL
CCV32 - 1	CCV32	26721	5.0999	-	7471B	PPB	26 May 2023 15:06:21	U	SMPL
CCB32 - 1	CCB32	-54	-0.0290	-	7471B	PPB	26 May 2023 15:08:37	U	SMPL
O2922-01 - 1	ORA-1618	2559	0.4715	-	7471B	PPB	26 May 2023 15:10:55	U	SMPL
O2923-01 - 1	OR-03-052423	2833	0.5240	-	7471B	PPB	26 May 2023 15:13:10	U	SMPL
O2937-01 - 1	AU-06-052423	5083	0.9550	-	7471B	PPB	26 May 2023 15:15:26	U	SMPL
O2945-01 - 1	SB20	6777	1.2795	-	7471B	PPB	26 May 2023 15:17:42	U	SMPL
O2945-03 - 1	SB17A	8396	1.5896	-	7471B	PPB	26 May 2023 15:19:58	U	SMPL
O2945-04 - 1	SB17B	2331	0.4279	-	7471B	PPB	26 May 2023 15:22:14	U	SMPL
O2945-06 - 1	SB23	3476	0.6472	-	7471B	PPB	26 May 2023 15:24:31	U	SMPL
O2916-01LX5 - 1		2814	0.5204	-	7471B	PPB	26 May 2023 15:31:22	U	SMPL
O2916-01A - 1		34511	6.5921	-	7471B	PPB	26 May 2023 15:33:40	U	SMPL
CCV33 - 1	CCV33	27020	5.1572	-	7471B	PPB	26 May 2023 15:35:56	U	SMPL
CCB33 - 1	CCB33	-76	-0.0332	-	7471B	PPB	26 May 2023 15:38:14	U	SMPL