

LB125681


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

4335E



A= 0.0000e+000

B= 2.3572e-004 slope

C= -4.1900e-002 intercept

Rho= 0.9988538

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.045	0.045	367	0.000	367					14
0.05	0.050					302	0				-2
0.20	0.200	0.227	0.027	1139	0.0 %	1139					3
2.50	2.500	2.441	-0.059	10532	0.0 %	10532					-5
5.00	5.000	5.161	0.161	22070	0.0 %	22070					2
7.50	7.500	7.149	-0.351	30505	0.0 %	30505					
10.0	10.000	10.179	0.179	43358	0.0 %	43358					

LB125681

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	367	-		0 SFAM01.1	PPB	5/26/2023 16:28	Std
	0.2 S01	1139	-		0.2 SFAM01.1	PPB	5/26/2023 16:30	Std
	2.5 S02	10532	-		2.5 SFAM01.1	PPB	5/26/2023 16:33	Std
	5 S03	22070	-		5 SFAM01.1	PPB	5/26/2023 16:35	Std
	7.5 S04	30505	-		7.5 SFAM01.1	PPB	5/26/2023 16:37	Std
	10 S05	43358	-		10 SFAM01.1	PPB	5/26/2023 16:42	Std
ICV096	ICV096	17638	4.1158 -		SFAM01.1	PPB	5/26/2023 16:48	SMPL
ICB096	ICB096	336	0.0373 -		SFAM01.1	PPB	5/26/2023 16:50	SMPL
CCV034	CCV034	22074	5.1615 -		SFAM01.1	PPB	5/26/2023 16:53	SMPL
CCB034	CCB034	338	0.0378 -		SFAM01.1	PPB	5/26/2023 16:55	SMPL
PB153089BL	PBW089	69	-0.0256 -		SFAM01.1	PPB	5/26/2023 16:59	SMPL
O2791-01	YBRR7	1622	0.3404 -		SFAM01.1	PPB	5/26/2023 17:06	SMPL
O2791-02	YBRR7D	1609	0.3374 -		SFAM01.1	PPB	5/26/2023 17:09	SMPL
O2791-03	YBRR7S	6273	1.4368 -		SFAM01.1	PPB	5/26/2023 17:11	SMPL
O2791-04	YBRR8	631	0.1068 -		SFAM01.1	PPB	5/26/2023 17:13	SMPL
CCV035	CCV035	20996	4.9074 -		SFAM01.1	PPB	5/26/2023 17:15	SMPL
CCB035	CCB035	335	0.0371 -		SFAM01.1	PPB	5/26/2023 17:18	SMPL