



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125696

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	05/26/2023	16:30
2	CAL2	1	Water	NA	NA	20.7	7.00	05/26/2023	16:31
3	CAL3	1	Water	NA	NA	20.6	10.02	05/26/2023	16:33
4	ICV	1	Water	NA	NA	20.6	6.98	05/26/2023	16:35
5	CCV1	1	Water	NA	NA	20.6	2.01	05/26/2023	16:38
6	O2958-04	1	Water	NA	NA	20.1	4.65	05/26/2023	16:40
7	O2958-04DUP	1	Water	NA	NA	20.2	4.66	05/26/2023	16:41
8	O2959-05	1	Water	NA	NA	23.3	5.52	05/26/2023	16:45
9	O2968-10	1	Water	NA	NA	24.4	6.94	05/26/2023	16:49
10	CCV2	1	Water	NA	NA	20.7	12.02	05/26/2023	16:50

WORKLIST(Hardcopy Internal Chain)

13 125696

WorkList Name : ph o2968

WorkList ID : 170386

Department : Wet-Chemistry

Date : 05-26-2023 14:38:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2958-04	427	Water	pH	Cool 4 deg C	PSEG03	D11	05/26/2023	9040C
O2959-05	60062	Water	pH	Cool 4 deg C	PSEG03	D11	05/26/2023	9040C
O2968-09	34281	Water	pH	Cool 4 deg C	PSEG03	D11	05/26/2023	9040C
O2968-10	34572	Water	pH	Cool 4 deg C	PSEG03	D11	05/26/2023	9040C

16

05/30/23

70
05/30/23

05/26/23 16:20

Date/Time

Raw Sample Received by:

JD (edc)

Raw Sample Relinquished by:

JD (edc)

Date/Time 05/26/23 18:00

Raw Sample Received by:

JD (edc)

Raw Sample Relinquished by:

JD (edc)