



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125697

Slope : 99.1

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	05/26/2023	16:30
2	CAL2	1	Water	NA	NA	20.7	7.00	05/26/2023	16:31
3	CAL3	1	Water	NA	NA	20.6	10.02	05/26/2023	16:33
4	ICV	1	Water	NA	NA	20.6	6.98	05/26/2023	16:35
5	CCV1	1	Water	NA	NA	20.6	2.01	05/26/2023	17:00
6	O2968-01	1	Solid	20.02	20	23.2	6.29	05/26/2023	17:10
7	O2968-01DUP	1	Solid	20.01	20	23.3	6.30	05/26/2023	17:11
8	O2968-12	1	Solid	20.01	20	23.2	5.24	05/26/2023	17:15
9	CCV2	1	Water	NA	NA	20.8	12.02	05/26/2023	17:20

WORKLIST(Hardcopy Internal Chain)

VB 125697

WorkList Name : ph ns 02968 WorkList ID : 170387 Department : Wet-Chemistry Date : 05-26-2023 14:39:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2968-01	34286	Solid	pH	Cool 4 deg C	PSEG03	D11	05/26/2023	9045D
O2968-12	34281	Solid	pH	Cool 4 deg C	PSEG03		05/26/2023	9045D

05/26/23 WLB

Date/Time
Raw Sample Received by: 20 (WC)
Raw Sample Relinquished by: 20 fwl

Date/Time
Raw Sample Received by: 20 (WC)
Raw Sample Relinquished by: 20 (WC)