



### Analytical Summary Report

Analysis Method: 1010B  
Parameter: Flash Point  
Run Number: LB125700  
Thermometer ID: FLASHPOINT

Reviewed By: RUBINA  
Supervisor Review By: Iwona  
Ambient Barometric Pressure (mmHg): 771.00  
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	05/30/2023	08:40
2	O2945-09		1	13	100.00	>212.0	>212.0	05/30/2023	09:10
3	O2945-09DUP		1	14	100.00	>212.0	>212.0	05/30/2023	09:40
4	O2945-11		1	12	100.00	>212.0	>212.0	05/30/2023	10:10
5	O2955-15		1	14	100.00	>212.0	>212.0	05/30/2023	10:40
6	O2958-04		1	13	100.00	>212.0	>212.0	05/30/2023	11:10
7	O2959-05		1	11	100.00	>212.0	>212.0	05/30/2023	11:40
8	O2968-10		1	14	100.00	>212.0	>212.0	05/30/2023	12:10

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

# WORKLIST(Hardcopy Internal Chain)

U0125700

WorkList Name : fp 5-30

WorkList ID : 170411

Department : Wet-Chemistry

Date : 05-30-2023 08:28:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2945-09	TWP03	Water	Flash Point	Cool 4 deg C	Ioui01	D11	05/24/2023	1010B
O2945-11	TWP04	Water	Flash Point	Cool 4 deg C	Ioui01	D11	05/24/2023	1010B
O2955-15	TWP05	Water	Flash Point	Cool 4 deg C	Ioui01	D11	05/25/2023	1010B
O2958-04	427	Water	Flash Point	Cool 4 deg C	PSEG03	D11	05/26/2023	1010B
O2959-05	60062	Water	Flash Point	Cool 4 deg C	PSEG03	D11	05/26/2023	1010B
O2968-10	34572	Water	Flash Point	Cool 4 deg C	PSEG03	D11	05/26/2023	1010B

Date/Time 05/30/2023 08:30  
 Raw Sample Received by: Rm Cure  
 Raw Sample Relinquished by: MCCOY

Date/Time 05/30/2023 12:45  
 Raw Sample Received by: Place  
 Raw Sample Relinquished by: Rm Cure