



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125701

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	05/30/2023	13:00
2	O2968-12		1	14	100.00	>212.0	>212.0	05/30/2023	13:30
3	O2968-12DUP		1	12	100.00	>212.0	>212.0	05/30/2023	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

L6125701

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp s-5-30

WorkList ID : 170430

Department : Wet-Chemistry

Date : 05-30-2023 12:46:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2968-12	34281	Solid	Flash Point	Cool 4 deg C	PSEG03	D11	05/26/2023	1010B

Date/Time 05/30/2023 12:50
Raw Sample Received by: RM Cava
Raw Sample Relinquished by: AP (OC)

Date/Time 05/30/2023 14:35
Raw Sample Received by: AP (OC)
Raw Sample Relinquished by: RM Cava