



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 5/31/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:00  
In Date: 05/30/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:10  
Out Date: 05/31/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB125705

| Lab ID   | Client SampleID           | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|---------------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| O2942-10 | OF-4-20230526             | 1      | 1.14           | 8.64         | 9.78                    | 7.12                      | 69.2    |                 |
| O2942-11 | OF-3-20230526             | 2      | 1.12           | 8.76         | 9.88                    | 7.51                      | 72.9    |                 |
| O2942-12 | O2942-11MS                | 3      | 1.12           | 8.76         | 9.88                    | 7.51                      | 72.9    |                 |
| O2942-13 | O2942-11MSD               | 4      | 1.12           | 8.76         | 9.88                    | 7.51                      | 72.9    |                 |
| O2942-14 | OF-2-20230526             | 5      | 1.16           | 8.80         | 9.96                    | 7.96                      | 77.3    |                 |
| O2942-15 | DUP-02-20230526           | 6      | 1.15           | 8.63         | 9.78                    | 7.17                      | 69.8    |                 |
| O2968-12 | 34281                     | 7      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | OIL SAMPLE      |
| O2972-01 | TP-11-0A                  | 8      | 1.16           | 8.82         | 9.98                    | 9.24                      | 91.6    |                 |
| O2972-03 | TP-11-0A-EPH-2            | 9      | 1.17           | 8.38         | 9.55                    | 8.6                       | 88.7    |                 |
| O2972-04 | TP-12-0A                  | 10     | 1.14           | 8.42         | 9.56                    | 8.48                      | 87.2    |                 |
| O2972-06 | TP-12-0A-EPH-2            | 11     | 1.19           | 8.78         | 9.97                    | 8.91                      | 87.9    |                 |
| O2974-01 | S27-WC-E07                | 12     | 1.15           | 8.81         | 9.96                    | 8.06                      | 78.4    |                 |
| O2974-02 | TP-WC-E08                 | 13     | 1.15           | 8.84         | 9.99                    | 8.87                      | 87.3    |                 |
| O2975-01 | 50007                     | 17     | 1.15           | 8.40         | 9.55                    | 9.4                       | 98.2    |                 |
| O2975-02 | 50008                     | 18     | 1.16           | 8.48         | 9.64                    | 9.52                      | 98.6    |                 |
| O2977-01 | FOUNDATION-TEST-PIT       | 14     | 1.18           | 8.73         | 9.91                    | 8.49                      | 83.7    |                 |
| O2977-03 | FOUNDATION-TEST-PIT-EPH-2 | 15     | 1.17           | 8.81         | 9.98                    | 8.59                      | 84.2    |                 |
| O2978-01 | WM6699AA                  | 16     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CONCRETE SAMPLE |
| O2979-01 | RBR044-VNJ224             | 19     | 1.18           | 8.64         | 9.82                    | 8.86                      | 88.9    |                 |
| O2980-01 | EO-01-05302023            | 20     | 1.17           | 8.70         | 9.87                    | 9.42                      | 94.8    |                 |
| O2980-02 | EO-01-05302023-EPH-2      | 21     | 1.19           | 8.50         | 9.69                    | 9.06                      | 92.6    |                 |
| O2981-01 | TP-3                      | 22     | 1.14           | 8.78         | 9.92                    | 9.27                      | 92.6    |                 |
| O2981-03 | TP-3-EPH-2                | 23     | 1.14           | 8.78         | 9.92                    | 9.27                      | 92.6    |                 |
| O2981-04 | TP-4                      | 24     | 1.18           | 8.78         | 9.96                    | 9.5                       | 94.8    |                 |
| O2981-06 | TP-4-EPH-2                | 25     | 1.18           | 8.78         | 9.96                    | 9.5                       | 94.8    |                 |
| O2981-07 | TP-5                      | 26     | 1.12           | 8.83         | 9.95                    | 9.12                      | 90.6    |                 |
| O2981-09 | TP-5-EPH-2                | 27     | 1.12           | 8.83         | 9.95                    | 9.12                      | 90.6    |                 |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$