

LB125706


7471B

INSTRUMENT ID (CV)

Linear

 μ Abs.:

59216



A= 0.0000e+000

B= 1.6618e-004

C= 1.7666e-002

Rho= 0.9995229

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.013	0.013	-26	0.000	-26					
0.2	0.200	0.221	0.021	1226	0.0 %	1226					11
2.5	2.500	2.357	-0.143	14080	0.0 %	14080					-6
5.0	5.000	5.072	0.072	30414	0.0 %	30414					1
7.5	7.500	7.678	0.178	46097	0.0 %	46097					2
10.0	10.000	9.858	-0.142	59216	0.0 %	59216					-1

LB125706 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 May 2023 14:43:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-26	-	0.0000	7471B	PPB	30 May 2023 15:36:33	S	Std
0.2 - 1	50.2	1226	-	0.2000	7471B	PPB	30 May 2023 15:38:50	S	Std
2.5 - 1	52.5	14080	-	2.5000	7471B	PPB	30 May 2023 15:41:08	S	Std
5.0 - 1	55	30414	-	5.0000	7471B	PPB	30 May 2023 15:43:25	S	Std
7.5 - 1	57.5	46097	-	7.5000	7471B	PPB	30 May 2023 15:45:42	S	Std
10.0 - 1	50	59216	-	10.0000	7471B	PPB	30 May 2023 15:48:00	S	Std
ICV97 - 1	ICV97	24307	4.0569	-7471B	PPB	30 May 2023 15:53:22	U	SMPL	
ICB97 - 1	ICB97	-243	-0.0227	-7471B	PPB	30 May 2023 15:55:39	U	SMPL	
CCV36 - 1	CCV36	30362	5.0632	-7471B	PPB	30 May 2023 15:57:56	U	SMPL	
CCB36 - 1	CCB36	-214	-0.0179	-7471B	PPB	30 May 2023 16:00:11	U	SMPL	
CRA - 1	CRA	1188	0.2151	-7471B	PPB	30 May 2023 16:02:28	U	SMPL	
HighStd - 1	HighStd	59047	9.8300	-7471B	PPB	30 May 2023 16:04:44	U	SMPL	
ChkStd - 1	ChkStd	42665	7.1076	-7471B	PPB	30 May 2023 16:06:59	U	SMPL	
PB153120BL - 1	PBS	-229	-0.0204	-7471B	PPB	30 May 2023 16:09:18	U	SMPL	
PB153120BS - 1	LCSS	24744	4.1296	-7471B	PPB	30 May 2023 16:11:37	U	SMPL	
O2953-02 - 1	PNNL-RH-524-1	11113	1.8644	-7471B	PPB	30 May 2023 16:13:53	U	SMPL	
O2953-02DUP - 1	PNNL-RH-524-1DUP	11605	1.9462	-7471B	PPB	30 May 2023 16:16:12	U	SMPL	
O2953-02MS - 1	PNNL-RH-524-1MS	34588	5.7654	-7471B	PPB	30 May 2023 16:23:52	U	SMPL	
O2953-02MSD - 1	PNNL-RH-524-1MSD	34527	5.7553	-7471B	PPB	30 May 2023 16:26:09	U	SMPL	
O2953-05 - 1	PNNL-RH-524-2	10867	1.8235	-7471B	PPB	30 May 2023 16:28:27	U	SMPL	
CCV37 - 1	CCV37	29834	4.9754	-7471B	PPB	30 May 2023 16:30:44	U	SMPL	
CCB37 - 1	CCB37	-148	-0.0069	-7471B	PPB	30 May 2023 16:33:01	U	SMPL	
O2955-01 - 1	SB27	7026	1.1852	-7471B	PPB	30 May 2023 16:35:19	U	SMPL	
O2955-02 - 1	DUP02	4296	0.7316	-7471B	PPB	30 May 2023 16:37:34	U	SMPL	
O2955-04 - 1	SB25	2051	0.3585	-7471B	PPB	30 May 2023 16:39:50	U	SMPL	
O2955-06 - 1	SB26	8053	1.3559	-7471B	PPB	30 May 2023 16:42:06	U	SMPL	
O2955-08 - 1	SB22	8398	1.4132	-7471B	PPB	30 May 2023 16:44:22	U	SMPL	
O2955-09 - 1	DUP03	9196	1.5458	-7471B	PPB	30 May 2023 16:46:38	U	SMPL	
O2955-11 - 1	SB21	2009	0.3515	-7471B	PPB	30 May 2023 16:48:55	U	SMPL	
O2957-01 - 1	TP-9-0A	412	0.0861	-7471B	PPB	30 May 2023 16:51:12	U	SMPL	
O2957-04 - 1	TP-10-0A	477	0.0969	-7471B	PPB	30 May 2023 16:53:29	U	SMPL	
O2958-01 - 1	ORA-0588	410	0.0858	-7471B	PPB	30 May 2023 16:55:45	U	SMPL	
CCV38 - 1	CCV38	29747	4.9610	-7471B	PPB	30 May 2023 16:58:01	U	SMPL	
CCB38 - 1	CCB38	-114	-0.0013	-7471B	PPB	30 May 2023 17:00:17	U	SMPL	
O2958-03 - 1	732-736	12954	2.1703	-7471B	PPB	30 May 2023 17:02:34	U	SMPL	
O2965-02 - 1	ETGI-338	2255	0.3924	-7471B	PPB	30 May 2023 17:04:50	U	SMPL	
O2968-12 - 1	34281	427	0.0886	-7471B	PPB	30 May 2023 17:07:07	U	SMPL	
O2971-01 - 1	SB16	1209	0.2186	-7471B	PPB	30 May 2023 17:09:23	U	SMPL	
O2971-03 - 1	SB15	1165	0.2113	-7471B	PPB	30 May 2023 17:11:39	U	SMPL	
O2971-05 - 1	SB12	885	0.1647	-7471B	PPB	30 May 2023 17:13:55	U	SMPL	
O2953-02LX5 - 1		1798	0.3165	-7471B	PPB	30 May 2023 17:16:12	U	SMPL	
O2953-02A - 1		33595	5.6004	-7471B	PPB	30 May 2023 17:18:29	U	SMPL	
CCV39 - 1	CCV39	29494	4.9189	-7471B	PPB	30 May 2023 17:20:46	U	SMPL	
CCB39 - 1	CCB39	-159	-0.0088	-7471B	PPB	30 May 2023 17:23:04	U	SMPL	