

LB125718


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

43737



A= 0.0000e+000

B= 2.2650e-004

C= 1.7639e-002

Rho= 0.9998082

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.013	0.013	-19	0.000	-19	0				2
0.2	0.200	0.205	0.005	826	0.0 %	826					-4
2.5	2.500	2.393	-0.107	10488	0.0 %	10488					2
5.0	5.000	5.096	0.096	22423	0.0 %	22423					1
7.5	7.500	7.568	0.068	33336	0.0 %	33336					-1
10.0	10.000	9.924	-0.076	43737	0.0 %	43737					

LB125718 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 May 2023 10:54:14

Sample ID	Extended ID	μ Abs	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-19	-	0.0000	7470A	PPB	31 May 2023 11:02:09	S	Std
0.2 - 1	SO.2	826	-	0.2000	7470A	PPB	31 May 2023 11:04:26	S	Std
2.5 - 1	SO.5	10488	-	2.5000	7470A	PPB	31 May 2023 11:06:43	S	Std
5.0 - 1	SO.5	22423	-	5.0000	7470A	PPB	31 May 2023 11:09:01	S	Std
7.5 - 1	SO.5	33336	-	7.5000	7470A	PPB	31 May 2023 11:11:17	S	Std
10.0 - 1	SO.5	43737	-	10.0000	7470A	PPB	31 May 2023 11:13:35	S	Std
ICV98 - 1	ICV98	18721	4.2579	-7470A	PPB	31 May 2023 11:16:21	U	SMPL	
ICB98 - 1	ICB98	-214	-0.0308	-7470A	PPB	31 May 2023 11:18:36	U	SMPL	
CCV40 - 1	CCV40	21387	4.8618	-7470A	PPB	31 May 2023 11:20:52	U	SMPL	
CCB40 - 1	CCB40	-176	-0.0222	-7470A	PPB	31 May 2023 11:23:07	U	SMPL	
CRA - 1	CRA	824	0.2043	-7470A	PPB	31 May 2023 11:25:24	U	SMPL	
HighStd - 1	HighStd	42810	9.7141	-7470A	PPB	31 May 2023 11:27:40	U	SMPL	
ChkStd - 1	ChkStd	31258	7.0976	-7470A	PPB	31 May 2023 11:29:55	U	SMPL	
PB153121BL - 1	PBW	-272	-0.0440	-7470A	PPB	31 May 2023 11:32:14	U	SMPL	
PB153121BS - 1	LCSW	19214	4.3696	-7470A	PPB	31 May 2023 11:37:13	U	SMPL	
O2945-02 - 1	WC20	-185	-0.0243	-7470A	PPB	31 May 2023 11:39:30	U	SMPL	
O2945-05 - 1	WC17	-10	0.0154	-7470A	PPB	31 May 2023 11:41:48	U	SMPL	
O2945-07 - 1	WC23	1	0.0179	-7470A	PPB	31 May 2023 11:44:05	U	SMPL	
O2953-03 - 1	PNNL-RH-524-1	73	0.0342	-7470A	PPB	31 May 2023 11:46:22	U	SMPL	
O2953-06 - 1	PNNL-RH-524-2	650	0.1649	-7470A	PPB	31 May 2023 11:48:38	U	SMPL	
CCV41 - 1	CCV41	21836	4.9635	-7470A	PPB	31 May 2023 11:50:53	U	SMPL	
CCB41 - 1	CCB41	-142	-0.0145	-7470A	PPB	31 May 2023 11:53:09	U	SMPL	
O2955-03 - 1	WC27	-9	0.0156	-7470A	PPB	31 May 2023 11:55:26	U	SMPL	
O2955-05 - 1	WC25	3	0.0183	-7470A	PPB	31 May 2023 11:57:41	U	SMPL	
O2955-07 - 1	WC26	18	0.0217	-7470A	PPB	31 May 2023 11:59:57	U	SMPL	
O2955-10 - 1	WC22	19	0.0219	-7470A	PPB	31 May 2023 12:02:13	U	SMPL	
O2955-12 - 1	WC21	27	0.0238	-7470A	PPB	31 May 2023 12:04:29	U	SMPL	
O2957-02 - 1	TP-9-0A	16	0.0213	-7470A	PPB	31 May 2023 12:06:45	U	SMPL	
O2957-05 - 1	TP-10-0A	51	0.0292	-7470A	PPB	31 May 2023 12:09:02	U	SMPL	
O2971-02 - 1	WC16	10	0.0199	-7470A	PPB	31 May 2023 12:11:19	U	SMPL	
O2971-04 - 1	WC15	14	0.0208	-7470A	PPB	31 May 2023 12:13:36	U	SMPL	
O2971-06 - 1	WC12	29	0.0242	-7470A	PPB	31 May 2023 12:15:52	U	SMPL	
CCV42 - 1	CCV42	21857	4.9683	-7470A	PPB	31 May 2023 12:18:08	U	SMPL	
CCB42 - 1	CCB42	-203	-0.0283	-7470A	PPB	31 May 2023 12:20:23	U	SMPL	
O2971-06DUP - 1	WC12DUP	-14	0.0145	-7470A	PPB	31 May 2023 12:22:42	U	SMPL	
O2971-06MS - 1	WC12MS	19427	4.4179	-7470A	PPB	31 May 2023 12:27:12	U	SMPL	
O2971-06MSD - 1	WC12MSD	17531	3.9884	-7470A	PPB	31 May 2023 12:29:28	U	SMPL	
PB153122BL - 1	PBW	-186	-0.0245	-7470A	PPB	31 May 2023 12:31:46	U	SMPL	
PB153122BS - 1	LCSW	17170	3.9066	-7470A	PPB	31 May 2023 12:39:00	U	SMPL	
O2955-13 - 1	TWP05	374	0.1024	-7470A	PPB	31 May 2023 12:48:33	U	SMPL	
O2955-13DUP - 1	TWP05DUP	285	0.0822	-7470A	PPB	31 May 2023 12:50:49	U	SMPL	
O2955-13MS - 1	TWP05MS	19055	4.3336	-7470A	PPB	31 May 2023 12:53:07	U	SMPL	
O2955-13MSD - 1	TWP05MSD	19001	4.3214	-7470A	PPB	31 May 2023 12:57:51	U	SMPL	
O2955-14 - 1	TWP05	-375	-0.0673	-7470A	PPB	31 May 2023 13:00:08	U	SMPL	
CCV43 - 1	CCV43	20315	4.6190	-7470A	PPB	31 May 2023 13:02:26	U	SMPL	
CCB43 - 1	CCB43	-200	-0.0277	-7470A	PPB	31 May 2023 13:04:42	U	SMPL	
O2955-16 - 1	DUP07	1438	0.3433	-7470A	PPB	31 May 2023 13:07:01	U	SMPL	
O2955-18 - 1	DUP07	-81	-0.0007	-7470A	PPB	31 May 2023 13:09:16	U	SMPL	
O2958-04 - 1	427	2808	0.6537	-7470A	PPB	31 May 2023 13:11:34	U	SMPL	
O2959-05 - 1	60062	13414	3.0559	-7470A	PPB	31 May 2023 13:13:50	U	SMPL	
O2971-07 - 1	FB01	-467	-0.0881	-7470A	PPB	31 May 2023 13:16:08	U	SMPL	
PB153087TB - 1	PB153087TB	-11	0.0151	-7470A	PPB	31 May 2023 13:18:28	U	SMPL	
O2971-06LX5 - 1		7	0.0192	-7470A	PPB	31 May 2023 13:20:45	U	SMPL	
O2971-06A - 1		19164	4.3583	-7470A	PPB	31 May 2023 13:23:02	U	SMPL	
O2955-13LX5 - 1		-351	-0.0619	-7470A	PPB	31 May 2023 13:25:19	U	SMPL	
O2955-13A - 1		18267	4.1551	-7470A	PPB	31 May 2023 13:27:37	U	SMPL	
CCV44 - 1	CCV44	22818	5.1859	-7470A	PPB	31 May 2023 13:29:53	U	SMPL	
CCB44 - 1	CCB44	-187	-0.0247	-7470A	PPB	31 May 2023 13:32:12	U	SMPL	