



TOTAL SUSPENDED SOLIDS - SM2540D

TEMP1 IN: 103 °C 05/29/2023 14:00 TEMP1 OUT: 103 °C 05/29/2023 15:00
TEMP2 IN: 104 °C 05/29/2023 15:30 TEMP2 OUT: 104 °C 05/29/2023 16:30
TEMP3 IN: 103 °C 05/30/2023 10:00 TEMP3 OUT: 104 °C 05/30/2023 11:30
TEMP4 IN: 104 °C 05/30/2023 12:00 TEMP4 OUT: 103 °C 05/30/2023 13:30

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 05/29/2023

Run Number: LB125725

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB125725BL	LB125725BL	1.3598	1.3598	100	1.3599	1.3599	1.3599	0.0001	1
2	LB125725BS	LB125725BS	1.5843	1.5843	100	1.6375	1.6375	1.6375	0.0532	532
3	O2945-09	TWP03	1.4004	1.4004	200	1.9253	1.9253	1.9253	0.5249	2624.5
4	O2945-11	TWP04	1.4087	1.4087	500	1.9854	1.9854	1.9854	0.5767	1153.4
5	O2945-11DUP	TWP04DUP	1.4009	1.4009	500	1.9779	1.9779	1.9779	0.5770	1154
6	O2952-01	001-WILLETS-PT-BLVD (MAY)	1.4147	1.4147	50	1.4201	1.4201	1.4201	0.0054	108
7	O2952-02	002-35TH-AVE (MAY)	1.4062	1.4062	50	1.4111	1.4111	1.4111	0.0049	98
8	O2955-15	TWP05	1.4055	1.4055	300	2.2457	2.2457	2.2457	0.8402	2800.7
9	O2969-01	001A- (JAN-JUNE)	1.3981	1.3981	50	1.4065	1.4065	1.4065	0.0084	168
10	O2969-02	002A- (JAN-JUNE)	1.4068	1.4068	50	1.4167	1.4167	1.4167	0.0099	198

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

5/12/23

WorkList Name : tss o2969

WorkList ID : 170400

Department : Wet-Chemistry

Date : 05-30-2023 08:12:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2945-09	TWP03	Water	TSS	Cool 4 deg C	loui01	D11	05/24/2023	SM2540 D
O2945-11	TWP04	Water	TSS	Cool 4 deg C	loui01	D11	05/24/2023	SM2540 D
O2952-01	001-WILLETTS-PT-BLVD(MAY)	Water	TSS	Cool 4 deg C	TULL01	D11	05/25/2023	SM2540 D
O2952-02	002-35TH-AVE(MAY)	Water	TSS	Cool 4 deg C	TULL01	D11	05/25/2023	SM2540 D
O2955-15	TWP05	Water	TSS	Cool 4 deg C	loui01	D11	05/25/2023	SM2540 D
O2969-01	001A-(JAN-JUNE)	Water	TSS	Cool 4 deg C	WILL06	D11	05/24/2023	SM2540 D
O2969-02	002-(JAN-JUNE)	Water	TSS	Cool 4 deg C	WILL06	D11	05/24/2023	SM2540 D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

05/30/23 08:15
JDC
JDC

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

05/30/23 21:00
JDC
JDC