

LB125730


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

59734



A= 0.0000e+000

B= 1.6793e-004

C= -1.0994e-002

Rho= 0.9997414

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.006	0.006	101	0.000	101					
0.2	0.200	0.216	0.016	1352	0.0 %	1352					
2.5	2.500	2.411	-0.089	14420	0.0 %	14420					
5.0	5.000	5.151	0.151	30740	0.0 %	30740					
7.5	7.500	7.396	-0.104	44111	0.0 %	44111					
10.0	10.000	10.020	0.020	59734	0.0 %	59734					

101
1352
14420
30740
44111
59734

LB125730 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 May 2023 15:14:30

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	101	-	0.0000	7471B	PPB	31 May 2023 15:51:22	S	Std
0.2 - 1	SD	1352	-	0.2000	7471B	PPB	31 May 2023 15:53:39	S	Std
2.5 - 1	SD	14420	-	2.5000	7471B	PPB	31 May 2023 15:55:57	S	Std
5.0 - 1	SD	30740	-	5.0000	7471B	PPB	31 May 2023 15:58:14	S	Std
7.5 - 1	SD	44111	-	7.5000	7471B	PPB	31 May 2023 16:00:32	S	Std
10.0 - 1	SD	59734	-	10.0000	7471B	PPB	31 May 2023 16:02:50	S	Std
ICV99 - 1	ICV99	25911	4.3402	-	7471B	PPB	31 May 2023 16:05:39	U	SMPL
ICB99 - 1	ICB99	-78	-0.0241	-	7471B	PPB	31 May 2023 16:07:55	U	SMPL
CCV45 - 1	CCV45	31069	5.2063	-	7471B	PPB	31 May 2023 16:10:11	U	SMPL
CCB45 - 1	CCB45	-102	-0.0281	-	7471B	PPB	31 May 2023 16:12:27	U	SMPL
CRA - 1	CRA	1284	0.2046	-	7471B	PPB	31 May 2023 16:14:44	U	SMPL
HighStd - 1	HighStd	58101	9.7457	-	7471B	PPB	31 May 2023 16:17:00	U	SMPL
ChkStd - 1	ChkStd	40723	6.8275	-	7471B	PPB	31 May 2023 16:19:16	U	SMPL
PB153146BL - 1	PBS	-211	-0.0464	-	7471B	PPB	31 May 2023 16:21:35	U	SMPL
PB153146BS - 1	LCSS	23545	3.9428	-	7471B	PPB	31 May 2023 16:26:29	U	SMPL
O2942-10 - 1	OF-4-20230526	4436	0.7339	-	7471B	PPB	31 May 2023 16:28:45	U	SMPL
O2942-11 - 1	OF-3-20230526	1697	0.2740	-	7471B	PPB	31 May 2023 16:35:57	U	SMPL
O2942-11DUP - 1	OF-3-20230526DUP	1551	0.2495	-	7471B	PPB	31 May 2023 16:38:14	U	SMPL
O2942-12 - 1	O2942-11MS	25565	4.2821	-	7471B	PPB	31 May 2023 16:40:31	U	SMPL
O2942-13 - 1	O2942-11MSD	25126	4.2083	-	7471B	PPB	31 May 2023 16:42:47	U	SMPL
CCV46 - 1	CCV46	29917	5.0129	-	7471B	PPB	31 May 2023 16:45:04	U	SMPL
CCB46 - 1	CCB46	-45	-0.0186	-	7471B	PPB	31 May 2023 16:47:21	U	SMPL
O2942-14 - 1	OF-2-20230526	4406	0.7289	-	7471B	PPB	31 May 2023 16:49:39	U	SMPL
O2942-15 - 1	DUP-02-20230526	4098	0.6772	-	7471B	PPB	31 May 2023 16:51:55	U	SMPL
O2972-01 - 1	TP-11-0A	507	0.0741	-	7471B	PPB	31 May 2023 16:54:11	U	SMPL
O2972-04 - 1	TP-12-0A	4928	0.8167	-	7471B	PPB	31 May 2023 16:56:27	U	SMPL
O2977-01 - 1	FOUNDATION-TEST-PIT	4513	0.7469	-	7471B	PPB	31 May 2023 16:58:43	U	SMPL
O2978-01 - 1	WM6699AA	136	0.0118	-	7471B	PPB	31 May 2023 17:01:00	U	SMPL
O2979-01 - 1	RBR044-VNJ224	75132	12.6057	-	7471B	PPB	31 May 2023 17:03:17	U	SMPL
O2980-01 - 1	EO-01-05302023	4763	0.7888	-	7471B	PPB	31 May 2023 17:05:34	U	SMPL
O2981-01 - 1	TP-3	631	0.0950	-	7471B	PPB	31 May 2023 17:07:54	U	SMPL
O2981-04 - 1	TP-4	479	0.0894	-	7471B	PPB	31 May 2023 17:10:10	U	SMPL
CCV47 - 1	CCV47	30679	5.1408	-	7471B	PPB	31 May 2023 17:12:26	U	SMPL
CCB47 - 1	CCB47	-24	-0.0150	-	7471B	PPB	31 May 2023 17:14:41	U	SMPL
O2981-07 - 1	TP-5	1399	0.2239	-	7471B	PPB	31 May 2023 17:16:58	U	SMPL
O2942-11LX5 - 1		219	0.0258	-	7471B	PPB	31 May 2023 17:19:14	U	SMPL
O2942-11A - 1		25802	4.3218	-	7471B	PPB	31 May 2023 17:21:30	U	SMPL
O2979-01DLX2 - 1		34133	5.7209	-	7471B	PPB	31 May 2023 17:23:46	U	SMPL
CCV48 - 1	CCV48	30052	5.0355	-	7471B	PPB	31 May 2023 17:30:40	U	SMPL
CCB48 - 1	CCB48	-39	-0.0175	-	7471B	PPB	31 May 2023 17:32:56	U	SMPL