

LB125738


7470A

INSTRUMENT ID: Cyl

Linear

μ Abs.:

44696



A= 0.0000e+000

B= 2.2385e-004

C= 3.6998e-002

Rho= 0.9998912

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	0.036	0.036	-5	0.000	-5	0				7
0.2	0.200	0.214	0.014	792	0.0 %	792					0
2.5	2.500	2.509	0.009	11045	0.0 %	11045					0
5.0	5.000	4.881	-0.119	21640	0.0 %	21640					-2
7.5	7.500	7.517	0.017	33415	0.0 %	33415					0
10.0	10.000	10.042	0.042	44696	0.0 %	44696					0

LB125738 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Jun 2023 10:07:51

Sample ID		Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		50	-5	-	0.0000	7470A	PPB	01 Jun 2023 10:23:32	S	Std
0.2 - 1		50.2	792	-	0.2000	7470A	PPB	01 Jun 2023 10:25:50	S	Std
2.5 - 1		52.5	11045	-	2.5000	7470A	PPB	01 Jun 2023 10:28:07	S	Std
5.0 - 1		55	21640	-	5.0000	7470A	PPB	01 Jun 2023 10:30:24	S	Std
7.5 - 1		57.5	33415	-	7.5000	7470A	PPB	01 Jun 2023 10:32:41	S	Std
10.0 - 1		516	44696	-	10.0000	7470A	PPB	01 Jun 2023 10:34:58	S	Std
ICV400 - 1	ICV01	ICV100	18428	4.1621	-7470A	PPB		01 Jun 2023 10:37:52	U	SMPL
ICB400 - 1	ICB01	ICB100	-174	-0.0020	-7470A	PPB		01 Jun 2023 10:40:07	U	SMPL
CCV49 - 1		CCV49	21551	4.8612	-7470A	PPB		01 Jun 2023 10:42:23	U	SMPL
CCB49 - 1		CCB49	-162	0.0007	-7470A	PPB		01 Jun 2023 10:44:38	U	SMPL
CRA - 1		CRA	756	0.2062	-7470A	PPB		01 Jun 2023 10:46:55	U	SMPL
HighStd - 1		HighStd	43519	9.7788	-7470A	PPB		01 Jun 2023 10:49:10	U	SMPL
ChkStd - 1		ChkStd	30488	6.8618	-7470A	PPB		01 Jun 2023 10:51:26	U	SMPL
PB153143BL - 1		PBW	-228	-0.0140	-7470A	PPB		01 Jun 2023 10:53:43	U	SMPL
PB153143BS - 1		LCSW	15116	3.4207	-7470A	PPB		01 Jun 2023 10:58:47	U	SMPL
O2972-02 - 1		TP-11-0A	-158	0.0016	-7470A	PPB		01 Jun 2023 11:01:03	U	SMPL
O2972-05 - 1		TP-12-0A	-26	0.0312	-7470A	PPB		01 Jun 2023 11:03:21	U	SMPL
O2976-02 - 1		TP-1	18	0.0410	-7470A	PPB		01 Jun 2023 11:05:38	U	SMPL
O2976-05 - 1		TP-2	11	0.0395	-7470A	PPB		01 Jun 2023 11:07:55	U	SMPL
O2977-02 - 1		FOUNDATION-TEST-PIT	10	0.0392	-7470A	PPB		01 Jun 2023 11:10:11	U	SMPL
CCV50 - 1		CCV50	22563	5.0878	-7470A	PPB		01 Jun 2023 11:15:10	U	SMPL
CCB50 - 1		CCB50	-141	0.0054	-7470A	PPB		01 Jun 2023 11:17:25	U	SMPL
O2978-02 - 1		WM6699AA	-37	0.0287	-7470A	PPB		01 Jun 2023 11:19:42	U	SMPL
O2979-02 - 1		RBR044-VNJ224	42	0.0464	-7470A	PPB		01 Jun 2023 11:21:58	U	SMPL
O2981-02 - 1		TP-3	5	0.0381	-7470A	PPB		01 Jun 2023 11:24:14	U	SMPL
O2981-05 - 1		TP-4	1	0.0372	-7470A	PPB		01 Jun 2023 11:26:31	U	SMPL
O2981-08 - 1		TP-5	-20	0.0325	-7470A	PPB		01 Jun 2023 11:28:47	U	SMPL
O2981-08DUP - 1		TP-5DUP	-46	0.0267	-7470A	PPB		01 Jun 2023 11:31:03	U	SMPL
O2981-08MS - 1		TP-5MS	15593	3.5275	-7470A	PPB		01 Jun 2023 11:33:20	U	SMPL
O2981-08MSD - 1		TP-5MSD	15523	3.5118	-7470A	PPB		01 Jun 2023 11:35:37	U	SMPL
CCV51 - 1		CCV51	21046	4.7482	-7470A	PPB		01 Jun 2023 11:40:37	U	SMPL
CCB51 - 1		CCB51	-163	0.0005	-7470A	PPB		01 Jun 2023 11:42:53	U	SMPL
O2981-08LX5 - 1			3	0.0377	-7470A	PPB		01 Jun 2023 11:45:10	U	SMPL
O2981-08A - 1			17351	3.9210	-7470A	PPB		01 Jun 2023 11:47:26	U	SMPL
PB153118TB - 1		PB153118TB	-139	0.0059	-7470A	PPB		01 Jun 2023 11:49:41	U	SMPL
CCV52 - 1		CCV52	20953	4.7274	-7470A	PPB		01 Jun 2023 11:51:58	U	SMPL
CCB52 - 1		CCB52	-145	0.0045	-7470A	PPB		01 Jun 2023 11:54:14	U	SMPL