



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125740

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	06/01/2023	09:00
2	CAL2	1	Water	NA	NA	20.6	7.00	06/01/2023	09:01
3	CAL3	1	Water	NA	NA	20.4	10.02	06/01/2023	09:05
4	ICV	1	Water	NA	NA	20.5	6.99	06/01/2023	09:10
5	CCV1	1	Water	NA	NA	20.6	2.01	06/01/2023	10:00
6	O2997-01	1	Water	NA	NA	20.6	6.37	06/01/2023	10:02
7	O2997-01DUP	1	Water	NA	NA	20.7	6.38	06/01/2023	10:03
8	O2997-02	1	Water	NA	NA	20.8	6.73	06/01/2023	10:10
9	CCV2	1	Water	NA	NA	20.6	12.02	06/01/2023	10:15

VB 125740

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O2997

WorkList ID : 170472

Department : Wet-Chemistry

Date : 06-01-2023 09:43:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2997-01	DMA-1103	Water	pH	Cool 4 deg C	ATCE02	D11	05/26/2023	9040C
O2997-02	DMA-1105	Water	pH	Cool 4 deg C	ATCE02	D11	05/26/2023	9040C

Date/Time 06/01/23 09:50
Raw Sample Received by: JPLCC1
Raw Sample Relinquished by: AP sm

Date/Time 06/01/23 12:30
Raw Sample Received by: AP sm
Raw Sample Relinquished by: JPLCC1