



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/2/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 16:30
In Date: 06/01/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:19
Out Date: 06/02/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125745

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3002-01	SB05A	1	1.12	8.75	9.87	8.44	83.7	
O3002-02	SB05B	2	1.16	8.44	9.6	7.94	80.3	
O3002-03	WC05	3	1.15	8.81	9.96	8.71	85.8	
O3002-04	SB04	4	1.13	8.48	9.61	8.38	85.5	
O3002-05	WC04	5	1.18	8.73	9.91	8.47	83.5	
O3002-06	SB03	6	1.15	8.83	9.98	8.5	83.2	
O3002-07	WC03	7	1.15	8.40	9.55	8.16	83.5	
O3002-08	SB02	8	1.19	8.59	9.78	8.55	85.7	
O3002-09	WC02	9	1.12	8.42	9.54	8.49	87.5	
O3004-01	TR-04-060123	18	1.15	8.52	9.67	9.05	92.7	
O3004-02	TR-04-060123-EPH-2	19	1.17	8.55	9.72	8.72	88.3	
O3006-01	EO-3-060123	20	1.15	8.81	9.96	9.18	91.1	
O3006-02	EO-3-060123-EPH-2	21	1.16	8.81	9.97	9.00	89.0	
O3008-01	TP-18-0A	10	1.15	8.80	9.95	8.49	83.4	
O3008-03	TP-18-0A-EPH-2	11	1.19	8.74	9.93	8.35	81.9	
O3008-04	TP-19-0A	12	1.12	8.70	9.82	8.85	88.9	
O3008-06	TP-19-0A-EPH-2	13	1.18	8.63	9.81	8.88	89.2	
O3009-01	WC-1	14	1.11	8.46	9.57	8.83	91.3	
O3009-02	WC-1-EPH-2	15	1.16	8.62	9.78	9.00	91.0	
O3010-01	CONC-WC-1	16	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3011-01	CO-115968	17	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O3014-01	WC4-(6-12)	22	1.15	8.78	9.93	9.46	94.6	
O3015-01	WC9-(12-17)	23	1.13	8.74	9.87	9.49	95.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$