



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125749

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/01/2023	14:50
2	O2997-01		1	14	100.00	>212.0	>212.0	06/01/2023	15:20
3	O2997-01DUP		1	13	100.00	>212.0	>212.0	06/01/2023	15:50
4	O2997-02		1	12	100.00	>212.0	>212.0	06/01/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66125749

WorkList Name : fp5-31

WorkList ID : 170443

Department : Wet-Chemistry

Date : 05-31-2023 08:22:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2997-01C1	DMA-1103	Water	Flash Point	Cool 4 deg C	ATCE02	D11	05/26/2023	1010B
O2997-02C1	DMA-1105	Water	Flash Point	Cool 4 deg C	ATCE02	D11	05/26/2023	1010B

Date/Time 06/01/2023 14:45
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: JH(CDC)

Date/Time 06/01/2023 16:20
 Raw Sample Received by: JH(CDC)
 Raw Sample Relinquished by: RM (W)