

LB125758


7471B

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

53844



A= 0.0000e+000

B= 1.8317e-004

C= 6.0736e-002

Rho= 0.9996848

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(V/D)
0.0	0.000	0.061	0.061	1	0.000	1					1
0.2	0.200	0.217	0.017	855	0.0 %	855					8
2.5	2.500	2.348	-0.152	12486	0.0 %	12486					-6
5.0	5.000	5.015	0.015	27047	0.0 %	27047					0
7.5	7.500	7.635	0.135	41351	0.0 %	41351					2
10.0	10.000	9.924	-0.076	53844	0.0 %	53844					-1

LB125753 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Jun 2023 15:51:27

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.0000	7471B	PPB	01 Jun 2023 15:53:43	S	Std
0.2 - 1	50.2	855	-	0.2000	7471B	PPB	01 Jun 2023 15:56:01	S	Std
2.5 - 1	52.5	12486	-	2.5000	7471B	PPB	01 Jun 2023 15:58:18	S	Std
5.0 - 1	55	27047	-	5.0000	7471B	PPB	01 Jun 2023 16:00:35	S	Std
7.5 - 1	57.5	41351	-	7.5000	7471B	PPB	01 Jun 2023 16:02:52	S	Std
10.0 - 1	510	53844	-	10.0000	7471B	PPB	01 Jun 2023 16:07:45	S	Std
ICV101 - 1	ICV101	21629	4.0226	-	7471B	PPB	01 Jun 2023 16:10:32	U	SMPL
ICB101 - 1	ICB101	-161	0.0312	-	7471B	PPB	01 Jun 2023 16:12:48	U	SMPL
CCV53 - 1	CCV53	25740	4.7757	-	7471B	PPB	01 Jun 2023 16:15:06	U	SMPL
CCB53 - 1	CCB53	-163	0.0309	-	7471B	PPB	01 Jun 2023 16:17:21	U	SMPL
CRA - 1	CRA	797	0.2067	-	7471B	PPB	01 Jun 2023 16:19:39	U	SMPL
HighStd - 1	HighStd	51872	9.5624	-	7471B	PPB	01 Jun 2023 16:21:54	U	SMPL
ChkStd - 1	ChkStd	35981	6.6516	-	7471B	PPB	01 Jun 2023 16:24:10	U	SMPL
PB153163BL - 1	PBS	-228	0.0190	-	7471B	PPB	01 Jun 2023 16:26:28	U	SMPL
PB153163BS - 1	LCSS	22505	4.1831	-	7471B	PPB	01 Jun 2023 16:28:47	U	SMPL
O2976-01 - 1	TP-1	2459	0.5112	-	7471B	PPB	01 Jun 2023 16:37:00	U	SMPL
O2976-01DUP - 1	TP-1DUP	1945	0.4170	-	7471B	PPB	01 Jun 2023 16:39:17	U	SMPL
O2976-01MS - 1	TP-1MS	28258	5.2369	-	7471B	PPB	01 Jun 2023 16:41:34	U	SMPL
O2976-01MSD - 1	TP-1MSD	28838	5.3431	-	7471B	PPB	01 Jun 2023 16:43:52	U	SMPL
O2976-04 - 1	TP-2	881	0.2221	-	7471B	PPB	01 Jun 2023 16:46:09	U	SMPL
CCV54 - 1	CCV54	27279	5.0576	-	7471B	PPB	01 Jun 2023 16:51:11	U	SMPL
CCB54 - 1	CCB54	-146	0.0340	-	7471B	PPB	01 Jun 2023 16:53:26	U	SMPL
O2983-01 - 1	SB18	1199	0.2804	-	7471B	PPB	01 Jun 2023 16:55:44	U	SMPL
O2983-03 - 1	SB19	1186	0.2780	-	7471B	PPB	01 Jun 2023 16:57:59	U	SMPL
O2983-05 - 1	SB24A	4003	0.7940	-	7471B	PPB	01 Jun 2023 17:00:15	U	SMPL
O2983-06 - 1	SB24B	7954	1.5177	-	7471B	PPB	01 Jun 2023 17:02:31	U	SMPL
O2983-08 - 1	SB07	8294	1.5800	-	7471B	PPB	01 Jun 2023 17:04:47	U	SMPL
O2985-01 - 1	PLYMOUTH-ENG-001	16682	3.1165	-	7471B	PPB	01 Jun 2023 17:07:05	U	SMPL
O2985-02 - 1	PLYMOUTH-XRDS-001	19431	3.6200	-	7471B	PPB	01 Jun 2023 17:09:21	U	SMPL
O2986-01 - 1	TP-13-0A	2988	0.6081	-	7471B	PPB	01 Jun 2023 17:11:40	U	SMPL
O2986-04 - 1	TP-14-0A	1012	0.2461	-	7471B	PPB	01 Jun 2023 17:13:58	U	SMPL
O2986-07 - 1	TP-15-0A	3038	0.6172	-	7471B	PPB	01 Jun 2023 17:16:14	U	SMPL
CCV55 - 1	CCV55	28988	5.3706	-	7471B	PPB	01 Jun 2023 17:27:04	U	SMPL
CCB55 - 1	CCB55	-207	0.0228	-	7471B	PPB	01 Jun 2023 17:29:19	U	SMPL
O2988-01 - 1	ORA-1644	1818	0.3937	-	7471B	PPB	01 Jun 2023 17:31:38	U	SMPL
O2998-01 - 1	PAT-2-053123	3045	0.6185	-	7471B	PPB	01 Jun 2023 17:33:54	U	SMPL
O3001-01 - 1	TP16-0A	679	0.1851	-	7471B	PPB	01 Jun 2023 17:36:09	U	SMPL
O3001-04 - 1	TP17-0A	331	0.1214	-	7471B	PPB	01 Jun 2023 17:38:25	U	SMPL
O2976-01LX5 - 1		342	0.1234	-	7471B	PPB	01 Jun 2023 17:40:41	U	SMPL
O2976-01A - 1		26093	4.8403	-	7471B	PPB	01 Jun 2023 17:42:57	U	SMPL
CCV56 - 1	CCV56	26201	4.8601	-	7471B	PPB	01 Jun 2023 17:45:15	U	SMPL
CCB56 - 1	CCB56	-370	-0.0070	-	7471B	PPB	01 Jun 2023 17:47:35	U	SMPL