

LB125759


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4388E



A= 0.0000e+000

B= 2.2388e-004

C= -5.8228e-002

Rho= 0.9990253

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	0.012	0.012	312	0.000	312					
0.2	0.200	0.197	-0.003	1139	0.0 %	1139					-2
2.5	2.500	2.374	-0.126	10865	0.0 %	10865					-5
5.0	5.000	5.057	0.057	22849	0.0 %	22849					1
7.5	7.500	7.793	0.293	35069	0.0 %	35069					4
10.0	10.000	9.767	-0.233	43888	0.0 %	43888					-2

LB125759 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Jun 2023 09:06:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	312	-	0.0000	7470A	PPB	02 Jun 2023 09:31:42	S	Std
0.2 - 1	50.2	1139	-	0.2000	7470A	PPB	02 Jun 2023 09:34:00	S	Std
2.5 - 1	52.5	10865	-	2.5000	7470A	PPB	02 Jun 2023 09:36:17	S	Std
5.0 - 1	55	22849	-	5.0000	7470A	PPB	02 Jun 2023 09:38:34	S	Std
7.5 - 1	57.5	35069	-	7.5000	7470A	PPB	02 Jun 2023 09:40:51	S	Std
10.0 - 1	510	43888	-	10.0000	7470A	PPB	02 Jun 2023 09:45:46	S	Std
ICV102 - 1	ICV102	19777	4.3694	-	7470A	PPB	02 Jun 2023 09:48:28	U	SMPL
ICB102 - 1	ICB102	-118	-0.0846	-	7470A	PPB	02 Jun 2023 09:50:43	U	SMPL
CCV57 - 1	CCV57	22057	4.8798	-	7470A	PPB	02 Jun 2023 09:52:59	U	SMPL
CCB57 - 1	CCB57	-103	-0.0813	-	7470A	PPB	02 Jun 2023 09:55:14	U	SMPL
CRA - 1	CRA	1013	0.1686	-	7470A	PPB	02 Jun 2023 10:02:34	U	SMPL
HighStd - 1	HighStd	41380	9.2058	-	7470A	PPB	02 Jun 2023 10:04:49	U	SMPL
ChkStd - 1	ChkStd	30837	6.8455	-	7470A	PPB	02 Jun 2023 10:07:04	U	SMPL
PB153166BL - 1	PBW	-169	-0.0961	-	7470A	PPB	02 Jun 2023 10:09:22	U	SMPL
PB153166BS - 1	LCSW	19304	4.2635	-	7470A	PPB	02 Jun 2023 10:11:40	U	SMPL
O2983-02 - 1	WC18	-101	-0.0808	-	7470A	PPB	02 Jun 2023 10:13:57	U	SMPL
O2983-04 - 1	WC19	-26	-0.0640	-	7470A	PPB	02 Jun 2023 10:16:15	U	SMPL
O2983-07 - 1	WC24	13	-0.0553	-	7470A	PPB	02 Jun 2023 10:18:32	U	SMPL
O2983-09 - 1	WC07	42	-0.0488	-	7470A	PPB	02 Jun 2023 10:20:49	U	SMPL
O2986-02 - 1	TP-13-0A	41	-0.0490	-	7470A	PPB	02 Jun 2023 10:23:04	U	SMPL
CCV58 - 1	CCV58	20388	4.5062	-	7470A	PPB	02 Jun 2023 10:25:20	U	SMPL
CCB58 - 1	CCB58	-102	-0.0811	-	7470A	PPB	02 Jun 2023 10:27:36	U	SMPL
O2986-05 - 1	TP-14-0A	-4	-0.0591	-	7470A	PPB	02 Jun 2023 10:29:52	U	SMPL
O2986-08 - 1	TP-15-0A	37	-0.0499	-	7470A	PPB	02 Jun 2023 10:32:08	U	SMPL
O3001-02 - 1	TP16-0A	-13	-0.0611	-	7470A	PPB	02 Jun 2023 10:34:25	U	SMPL
O3001-05 - 1	TP17-0A	26	-0.0524	-	7470A	PPB	02 Jun 2023 10:36:42	U	SMPL
O3002-03 - 1	WC05	13	-0.0553	-	7470A	PPB	02 Jun 2023 10:38:58	U	SMPL
O3002-05 - 1	WC04	16	-0.0546	-	7470A	PPB	02 Jun 2023 10:41:16	U	SMPL
O3002-07 - 1	WC03	-12	-0.0609	-	7470A	PPB	02 Jun 2023 10:43:32	U	SMPL
O3002-09 - 1	WC02	-26	-0.0640	-	7470A	PPB	02 Jun 2023 10:45:50	U	SMPL
O3002-09DUP - 1	WC02DUP	-89	-0.0782	-	7470A	PPB	02 Jun 2023 10:48:08	U	SMPL
O3002-09MS - 1	WC02MS	16868	3.7181	-	7470A	PPB	02 Jun 2023 10:50:24	U	SMPL
CCV59 - 1	CCV59	23201	5.1360	-	7470A	PPB	02 Jun 2023 11:01:09	U	SMPL
CCB59 - 1	CCB59	-183	-0.0992	-	7470A	PPB	02 Jun 2023 11:03:24	U	SMPL
O3002-09MSD - 1	WC02MSD	16065	3.5384	-	7470A	PPB	02 Jun 2023 11:05:42	U	SMPL
PB153167BL - 1	PBW	-132	-0.0878	-	7470A	PPB	02 Jun 2023 11:07:58	U	SMPL
PB153167BS - 1	LCSW	16559	3.6490	-	7470A	PPB	02 Jun 2023 11:10:14	U	SMPL
O3002-10 - 1	MW-48	-121	-0.0853	-	7470A	PPB	02 Jun 2023 11:12:30	U	SMPL
O3002-10DUP - 1	MW-48DUP	-22	-0.0632	-	7470A	PPB	02 Jun 2023 11:14:47	U	SMPL
O3002-10MS - 1	MW-48MS	15312	3.3698	-	7470A	PPB	02 Jun 2023 11:17:03	U	SMPL
O3002-10MSD - 1	MW-48MSD	16219	3.5728	-	7470A	PPB	02 Jun 2023 11:19:21	U	SMPL
O3002-11 - 1	MW-48	-118	-0.0846	-	7470A	PPB	02 Jun 2023 11:21:39	U	SMPL
PB153150TB - 1	PB153150TB	-43	-0.0679	-	7470A	PPB	02 Jun 2023 11:23:57	U	SMPL
CCV60 - 1	CCV60	21722	4.8048	-	7470A	PPB	02 Jun 2023 11:26:13	U	SMPL
CCB60 - 1	CCB60	-164	-0.0949	-	7470A	PPB	02 Jun 2023 11:28:29	U	SMPL
O3002-09LX5 - 1		-9	-0.0602	-	7470A	PPB	02 Jun 2023 11:30:45	U	SMPL
O3002-09A - 1		17579	3.8773	-	7470A	PPB	02 Jun 2023 11:33:01	U	SMPL
O3002-10LX5 - 1		-107	-0.0822	-	7470A	PPB	02 Jun 2023 11:35:17	U	SMPL
O3002-10A - 1		17128	3.7763	-	7470A	PPB	02 Jun 2023 11:37:34	U	SMPL
CCV61 - 1	CCV61	21148	4.6763	-	7470A	PPB	02 Jun 2023 11:39:50	U	SMPL
CCB61 - 1	CCB61	-158	-0.0936	-	7470A	PPB	02 Jun 2023 11:42:07	U	SMPL