



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/5/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/02/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 06/03/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125763

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3013-03	MW-5B	1	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O3016-01	SB28	2	1.12	8.55	9.67	8.8	89.8	
O3016-02	WC28	3	1.12	8.59	9.71	8.76	88.9	
O3016-03	DUP04	4	1.19	8.72	9.91	8.92	88.6	
O3016-04	SB29	5	1.14	8.72	9.86	8.94	89.4	
O3016-05	WC29	6	1.13	8.85	9.98	8.82	86.9	
O3018-01	277	34	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3018-03	362	35	1.12	8.56	9.68	8.86	90.4	
O3018-05	348	36	1.17	8.60	9.77	8.48	85.0	
O3018-07	343	37	1.18	8.52	9.7	8.31	83.7	
O3018-09	72-11991	38	1.15	8.81	9.96	9.05	89.7	
O3018-11	278	39	1.12	8.45	9.57	8.35	85.6	
O3018-13	RT-4648	40	1.19	8.38	9.57	8.5	87.2	
O3020-01	P-060123RB01	7	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3020-02	P-060123RB02	8	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3020-03	P-060123RB03	9	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O3023-01	32415-32417	41	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O3026-01	VNJ-223	42	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3027-05	SVOC-GPC-BLANK	10	1.00	1.00	2.00	2.00	100.0	
O3027-06	PEST-GPC-BLANK	11	1.00	1.00	2.00	2.00	100.0	
O3027-07	PEST-GPC-BLANK-SPIKE	12	1.00	1.00	2.00	2.00	100.0	
O3027-08	PCB-GPC-BLANK	13	1.00	1.00	2.00	2.00	100.0	
O3027-09	PCB-GPC-BLANK-SPIKE	14	1.00	1.00	2.00	2.00	100.0	
O3027-10	SVOC-GPC2-BLANK	15	1.00	1.00	2.00	2.00	100.0	
O3027-11	PEST-GPC2-BLANK	16	1.00	1.00	2.00	2.00	100.0	
O3027-12	PEST-GPC2-BLANK-SPIKE	17	1.00	1.00	2.00	2.00	100.0	
O3027-13	PCB-GCP2-BLANK	18	1.00	1.00	2.00	2.00	100.0	
O3027-14	PCB-GCP2-BLANK-SPIKE	19	1.00	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/5/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/02/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 06/03/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125763

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3029-01	TP-6	20	1.14	8.76	9.9	8.99	89.6	
O3029-03	TP-6-EPH-2	21	1.11	8.56	9.67	8.77	89.5	
O3029-04	TP-7	22	1.15	8.80	9.95	9.04	89.7	
O3029-06	TP-7-EPH-2	23	1.12	8.76	9.88	9.02	90.2	
O3029-07	TP-8	24	1.17	8.67	9.84	8.72	87.1	
O3029-09	TP-8-EPH-2	25	1.18	8.48	9.66	8.61	87.6	
O3029-10	TP-9	26	1.15	8.63	9.78	8.59	86.2	
O3029-12	TP-9-EPH-2	27	1.14	8.45	9.59	8.44	86.4	
O3029-13	TP-10	28	1.15	8.84	9.99	9.28	92.0	
O3029-15	TP-10-EPH-2	29	1.18	8.49	9.67	8.89	90.8	
O3032-01	TP-20-0A	30	1.18	8.72	9.9	9.24	92.4	
O3032-03	TP-20-0A-EPH-2	31	1.16	8.81	9.97	9.1	90.1	
O3032-04	TP-21-0A	32	1.16	8.60	9.76	8.78	88.6	
O3032-06	TP-21-0A-EPH-2	33	1.12	8.50	9.62	8.63	88.4	
O3036-01	PL-STONE-0602	43	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3036-03	PL-STONE-0602-EPH-2	44	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$