

LB125764


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41591



A= 0.0000e+000

B= 2.4063e-004

C= 1.1358e-002

Rho= 0.9999085

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	31	0.000	31					16
0.05	0.050	0.058	0.008	195	0.0 %	195	0				-10
0.2	0.200	0.180	-0.020	701	0.0 %	701					-2
2.5	2.500	2.455	-0.045	10155	0.0 %	10155					2
5.0	5.000	5.094	0.094	21121	0.0 %	21121					-1
7.5	7.500	7.425	-0.075	30809	0.0 %	30809					0
10.0	10.000	10.019	0.019	41591	0.0 %	41591					

LB125764 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Jun 2023 12:32:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	245.1	PPB	02 Jun 2023 12:53:12	S	Std
0.05 - 1	5005	195	-	0.0500	245.1	PPB	02 Jun 2023 12:55:30	S	Std
0.2 - 1	512	701	-	0.2000	245.1	PPB	02 Jun 2023 12:57:47	S	Std
2.5 - 1	525	10155	-	2.5000	245.1	PPB	02 Jun 2023 13:00:04	S	Std
5.0 - 1	55	21121	-	5.0000	245.1	PPB	02 Jun 2023 13:02:22	S	Std
7.5 - 1	575	30809	-	7.5000	245.1	PPB	02 Jun 2023 13:07:14	S	Std
10.0 - 1	510	41591	-	10.0000	245.1	PPB	02 Jun 2023 13:14:43	S	Std
ICV103 - 1	ICV103	17317	4.1783	-245.1		PPB	02 Jun 2023 13:17:27	U	SMPL
ICB103 - 1	ICB103	-210	-0.0392	-245.1		PPB	02 Jun 2023 13:19:44	U	SMPL
CCV62 - 1	CCV62	20777	5.0109	-245.1		PPB	02 Jun 2023 13:22:00	U	SMPL
CCB62 - 1	CCB62	-176	-0.0310	-245.1		PPB	02 Jun 2023 13:24:15	U	SMPL
CRA - 1	CRA	172	0.0527	-245.1		PPB	02 Jun 2023 13:26:32	U	SMPL
HighStd - 1	HighStd	41167	9.9173	-245.1		PPB	02 Jun 2023 13:28:48	U	SMPL
ChkStd - 1	ChkStd	28531	6.8767	-245.1		PPB	02 Jun 2023 13:31:03	U	SMPL
PB153174BL - 1	PBW	-365	-0.0765	-245.1		PPB	02 Jun 2023 13:33:24	U	SMPL
PB153174BS - 1	LCSW	16932	4.0857	-245.1		PPB	02 Jun 2023 13:35:42	U	SMPL
O2955-15 - 1	TWP05	1791	0.4423	-245.1		PPB	02 Jun 2023 13:42:29	U	SMPL
O2955-15DUP - 1	TWP05DUP	1657	0.4101	-245.1		PPB	02 Jun 2023 13:46:47	U	SMPL
O2955-15MS - 1	TWP05MS	19187	4.6283	-245.1		PPB	02 Jun 2023 13:49:04	U	SMPL
O2955-15MSD - 1	TWP05MSD	18302	4.4153	-245.1		PPB	02 Jun 2023 13:51:22	U	SMPL
CCV63 - 1	CCV63	20410	4.9226	-245.1		PPB	02 Jun 2023 13:53:45	U	SMPL
CCB63 - 1	CCB63	-125	-0.0187	-245.1		PPB	02 Jun 2023 13:56:02	U	SMPL
O2955-15LX5 - 1		303	0.0843	-245.1		PPB	02 Jun 2023 13:58:20	U	SMPL
O2955-15A - 1		17112	4.1290	-245.1		PPB	02 Jun 2023 14:00:36	U	SMPL
CCV64 - 1	CCV64	19993	4.8222	-245.1		PPB	02 Jun 2023 14:02:51	U	SMPL
CCB64 - 1	CCB64	-104	-0.0137	-245.1		PPB	02 Jun 2023 14:05:09	U	SMPL