

LB125769


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

59941



A= 0.0000e+000

B= 1.6527e-004

C= 1.1944e-002

Rho= 0.9995302

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0.001
0.0	0.000	0.067	0.067	334	0.000	334					1
0.2	0.200	0.190	-0.010	1075	0.0 %	1075					1
2.5	2.500	2.309	-0.191	13897	0.0 %	13897					1
5.0	5.000	5.161	0.161	31153	0.0 %	31153					3
7.5	7.500	7.556	0.056	45644	0.0 %	45644					1
10.0	10.000	9.918	-0.082	59941	0.0 %	59941					1

LB125769 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Jun 2023 15:23:21

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	334	-	0.0000	7471B	PPB	02 Jun 2023 15:56:34	S	Std
0.2 - 1	50.2	1075	-	0.2000	7471B	PPB	02 Jun 2023 15:58:53	S	Std
2.5 - 1	52.5	13897	-	2.5000	7471B	PPB	02 Jun 2023 16:01:10	S	Std
5.0 - 1	55	31153	-	5.0000	7471B	PPB	02 Jun 2023 16:03:27	S	Std
7.5 - 1	57.5	45644	-	7.5000	7471B	PPB	02 Jun 2023 16:05:44	S	Std
10.0 - 1	510	59941	-	10.0000	7471B	PPB	02 Jun 2023 16:10:36	S	Std
ICV104 - 1	ICV104	25540	4.2329	-	7471B	PPB	02 Jun 2023 16:13:50	U	SMPL
ICB104 - 1	ICB104	-167	-0.0157	-	7471B	PPB	02 Jun 2023 16:16:06	U	SMPL
CCV65 - 1	CCV65	29837	4.9431	-	7471B	PPB	02 Jun 2023 16:18:23	U	SMPL
CCB65 - 1	CCB65	-136	-0.0105	-	7471B	PPB	02 Jun 2023 16:20:38	U	SMPL
CRA - 1	CRA	902	0.1610	-	7471B	PPB	02 Jun 2023 16:22:54	U	SMPL
HighStd - 1	HighStd	59321	9.8159	-	7471B	PPB	02 Jun 2023 16:25:09	U	SMPL
ChkStd - 1	ChkStd	40725	6.7426	-	7471B	PPB	02 Jun 2023 16:27:25	U	SMPL
PB153197BL - 1	PBS	-198	-0.0208	-	7471B	PPB	02 Jun 2023 16:29:43	U	SMPL
PB153197BS - 1	LCSS	22403	3.7145	-	7471B	PPB	02 Jun 2023 16:32:01	U	SMPL
O3002-01 - 1	SB05A	20754	3.4420	-	7471B	PPB	02 Jun 2023 16:34:17	U	SMPL
O3002-02 - 1	SB05B	75371	12.4685	-	7471B	PPB	02 Jun 2023 16:36:36	U	SMPL
O3002-04 - 1	SB04	5239	0.8778	-	7471B	PPB	02 Jun 2023 16:38:54	U	SMPL
O3002-06 - 1	SB03	27309	4.5253	-	7471B	PPB	02 Jun 2023 16:41:14	U	SMPL
O3002-08 - 1	SB02	11782	1.9592	-	7471B	PPB	02 Jun 2023 16:43:30	U	SMPL
CCV66 - 1	CCV66	29738	4.9267	-	7471B	PPB	02 Jun 2023 16:45:47	U	SMPL
CCB66 - 1	CCB66	-155	-0.0137	-	7471B	PPB	02 Jun 2023 16:48:04	U	SMPL
O3004-01 - 1	TR-04-060123	4847	0.8130	-	7471B	PPB	02 Jun 2023 16:50:21	U	SMPL
O3006-01 - 1	EO-3-060123	2453	0.4174	-	7471B	PPB	02 Jun 2023 16:52:36	U	SMPL
O3008-01 - 1	TP-18-0A	3342	0.5643	-	7471B	PPB	02 Jun 2023 16:54:53	U	SMPL
O3008-04 - 1	TP-19-0A	407	0.0792	-	7471B	PPB	02 Jun 2023 16:57:09	U	SMPL
O3009-01 - 1	WC-1	8290	1.3820	-	7471B	PPB	02 Jun 2023 16:59:25	U	SMPL
O3009-01DUP - 1	WC-1DUP	8841	1.4731	-	7471B	PPB	02 Jun 2023 17:01:41	U	SMPL
O3009-01MS - 1	WC-1MS	29381	4.8677	-	7471B	PPB	02 Jun 2023 17:03:58	U	SMPL
O3009-01MSD - 1	WC-1MSD	32760	5.4262	-	7471B	PPB	02 Jun 2023 17:06:15	U	SMPL
O3016-01 - 1	SB28	472	0.0900	-	7471B	PPB	02 Jun 2023 17:08:34	U	SMPL
O3016-03 - 1	DUP04	735	0.1334	-	7471B	PPB	02 Jun 2023 17:10:52	U	SMPL
CCV67 - 1	CCV67	29567	4.8985	-	7471B	PPB	02 Jun 2023 17:13:08	U	SMPL
CCB67 - 1	CCB67	-155	-0.0137	-	7471B	PPB	02 Jun 2023 17:15:24	U	SMPL
O3016-04 - 1	SB29	452	0.0866	-	7471B	PPB	02 Jun 2023 17:17:42	U	SMPL
O3009-01LX5 - 1		1343	0.2339	-	7471B	PPB	02 Jun 2023 17:19:58	U	SMPL
O3009-01A - 1		30546	5.0603	-	7471B	PPB	02 Jun 2023 17:22:15	U	SMPL
O3002-02DLX2 - 1		35641	5.9023	-	7471B	PPB	02 Jun 2023 17:24:30	U	SMPL
CCV68 - 1	CCV68	29916	4.9562	-	7471B	PPB	02 Jun 2023 17:29:39	U	SMPL
CCB68 - 1	CCB68	-161	-0.0147	-	7471B	PPB	02 Jun 2023 17:31:55	U	SMPL