



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125778

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/04/2023	11:15
2	CAL2	1	Water	NA	NA	20.7	7.00	06/04/2023	11:18
3	CAL3	1	Water	NA	NA	20.7	10.02	06/05/2023	11:20
4	ICV	1	Water	NA	NA	20.8	6.99	06/04/2023	11:22
5	CCV1	1	Water	NA	NA	20.7	2.01	06/04/2023	11:25
6	O3019-01	1	Water	NA	NA	22.6	6.70	06/04/2023	11:30
7	O3019-01DUP	1	Water	NA	NA	22.7	6.71	06/04/2023	11:31
8	O3037-01	1	Water	NA	NA	21.4	6.10	06/04/2023	11:33
9	O3037-03	1	Water	NA	NA	21.8	6.84	06/04/2023	11:40
10	O3037-05	1	Water	NA	NA	21.2	6.94	06/04/2023	11:44
11	O3037-07	1	Water	NA	NA	20.7	7.20	06/04/2023	11:48
12	CCV2	1	Water	NA	NA	20.7	12.02	06/04/2023	11:50

WORKLIST(Hardcopy Internal Chain)

12578

WorkList Name : ph o3037 WorkList ID : 170529 Department : Wet-Chemistry Date : 06-04-2023 11:01:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3019-01	MH-289	Water	pH	Cool 4 deg C	PSEG03	D11	06/02/2023	9040C
O3037-01	MW-1	Water	pH	Cool 4 deg C	LOCK01	D11	06/01/2023	9040C
O3037-03	MW-3	Water	pH	Cool 4 deg C	LOCK01	D11	06/01/2023	9040C
O3037-05	MW-4	Water	pH	Cool 4 deg C	LOCK01	D11	06/01/2023	9040C
O3037-07	MW-2	Water	pH	Cool 4 deg C	LOCK01	D11	06/01/2023	9040C

Date/Time 06/04/23 11:10
 Raw Sample Received by: JWC
 Raw Sample Relinquished by: JWC

Date/Time 06/04/23 13:00
 Raw Sample Received by: JWC
 Raw Sample Relinquished by: JWC