



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 06/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 06/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125780

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3025-01	WASTE	1	1.14	8.49	9.63	8.41	85.6	
O3025-02	VOC	2	1.15	8.38	9.53	8.34	85.8	
O3025-03	1	3	1.16	8.69	9.85	8.51	84.6	
O3025-04	2	4	1.16	8.80	9.96	7.35	70.3	
O3025-05	3	5	1.13	8.67	9.8	8.47	84.7	
O3025-06	4	6	1.13	8.70	9.83	8.82	88.4	
O3025-07	5	7	1.18	8.76	9.94	8.72	86.1	
O3028-01	OR-02-060223	8	1.18	8.74	9.92	9.6	96.3	
O3028-02	OR-02-060223-EPH-2	9	1.18	8.74	9.92	9.6	96.3	
O3038-01	NB-07-060223	10	1.14	8.82	9.96	9.32	92.7	
O3038-02	NB-07-060223-EPH-2	11	1.14	8.82	9.96	9.32	92.7	
O3039-01	SB30A	12	1.12	8.84	9.96	7.93	77.0	
O3039-02	SB30B	13	1.14	8.81	9.95	8.43	82.7	
O3039-04	SB31A	14	1.13	8.53	9.66	8.2	82.9	
O3039-05	SB31B	15	1.13	8.62	9.75	8.72	88.1	
O3039-06	WC31	16	1.19	8.48	9.67	8.5	86.2	
O3043-01	MR-PIPE-A	17	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-02	MR-PIPE-B	18	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-03	MR-PIPE-C	19	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-04	MR-PIPE-D	20	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-05	MR-PIPE-E	21	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-06	MR-PIPE-F	22	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-07	MR-PIPE-G	23	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-08	MR-PIPE-H	24	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-09	MR-PIPE-I	25	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-10	MR-PIPE-J	26	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-11	MR-PIPE-K	27	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
O3043-12	MR-PIPE-L	28	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/6/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 06/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 06/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125780

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3045-01	PAT-1-060523	29	1.16	8.49	9.65	9.35	96.5	
O3045-02	PAT-1-060523-EPH-2	30	1.12	8.48	9.6	9.42	97.9	
O3049-01	HR-01-060523	35	1.18	8.68	9.86	9.47	95.5	
O3049-02	HR-01-060523-EPH-2	36	1.18	8.68	9.86	9.47	95.5	
O3049-03	HR-03-060523	37	1.12	8.85	9.97	9.5	94.7	
O3049-04	HR-03-060523-EPH-2	38	1.12	8.85	9.97	9.5	94.7	
O3050-01	TP-22-0A	31	1.16	8.48	9.64	8.84	90.6	
O3050-03	TP-22-0A-EPH-2	32	1.18	8.78	9.96	9.19	91.2	
O3050-04	TP-23-0A	33	1.18	8.80	9.98	9.13	90.3	
O3050-06	TP-23-0A-EPH-2	34	1.11	8.75	9.86	8.92	89.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$