



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 06/04/2023 15:00 TEMP1 OUT: 103 °C 06/05/2023 07:00
TEMP2 IN: 103 °C 06/05/2023 07:30 TEMP2 OUT: 103 °C 06/05/2023 09:00
TEMP3 IN: 540 °C 06/05/2023 09:30 TEMP3 OUT: 560 °C 06/05/2023 11:00
TEMP4 IN: 560 °C 06/05/2023 11:30 TEMP4 OUT: 560 °C 06/05/2023 13:00

Run Number: LB125785

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB125785BL	85.6941	85.6941	100	85.6941	85.6941	85.6941	85.6941	0.0000	0
2	O3019-01	94.8133	94.8133	100	94.8654	94.8654	94.8526	94.8526	0.0128	128

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

US 125785

WORKLIST(Hardcopy Internal Chain)

WorkList Name :
tvs o3019

WorkList ID :
170526

Department :
Wet-Chemistry

Date :
06-04-2023 06:29:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3019-01	MH-289	Water	TVS	Cool 4 deg C	PSEG03	D11	06/02/2023	SM2540 G

Date/Time
06/04/23 07:00

Raw Sample Received by:
[Signature]

Raw Sample Relinquished by:
[Signature]

Date/Time
06/04/23 16:00

Raw Sample Received by:
[Signature]

Raw Sample Relinquished by:
[Signature]