




Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB125788
Analysis Date: 06/05/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 06/05/2023
Extraction IN Time: 13:00
Extraction OUT Time: 13:30
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB125788BL	LB125788BL	WATER	1.3	1000	100	3.5624	3.5624	3.02	3.5625	3.5625	0.0001	0.1
2	LB125788BS	LB125788BS	WATER	1.3	1000	100	2.9874	2.9874	3.01	3.0041	3.0041	0.0167	16.7
3	LB125788BSD	LB125788BSD	WATER	1.3	1000	100	3.0254	3.0254	3.03	3.0418	3.0418	0.0164	16.4
4	O3016-08	B-45 (OW)	WATER	1.3	1000	100	3.0415	3.0415	3.05	3.0416	3.0416	0.0001	0.1
5	O3039-09	MW-34	WATER	1.6	1000	100	3.0638	3.0638	3.04	3.0641	3.0641	0.0003	0.3



Chemical Name	Chemical Lot #
HEXANE	W3010
pH Paper 0-14	M4909
Sodium Sulfate	EP2340
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.002</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1 :	<u>14:41</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>14:00</u>		
Bal Check Time:	<u>13:02</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1:	<u>15:30</u>
			Out Time1:	<u>14:40</u>		

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2 :	<u>16:31</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time2:	<u>16:00</u>		
Bal Check Time:	<u>17:15</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2:	<u>17:05</u>
			Out Time2:	<u>16:30</u>		

125788

WORKLIST(Hardcopy Internal Chain)

WorkList Name : non poplar o3039

WorkList ID : 170556

Department : Wet-Chemistry

Date : 06-05-2023 12:15:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3016-08	B-45(OW)	Water	Non-Polar Material	Conc H2SO4 to pH < 2	Ioui01	D11	06/01/2023	1664A
O3039-09	MW-34	Water	Non-Polar Material	Conc H2SO4 to pH < 2	Ioui01	D11	06/02/2023	1664A

Date/Time 06/05/23 12:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/05/23 16:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]