



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB125789
Analysis Date: 06/05/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 06/05/2023
Extraction IN Time: 13:00
Extraction OUT Time: 14:00
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB125789BL	LB125789BL	WATER	1.3	1000	100	3.0251	3.0251	3.02	3.0252	3.0252	0.0001	0.1
2	LB125789BS	LB125789BS	WATER	1.3	1000	100	3.1987	3.1987	3.01	3.2157	3.2157	0.0170	17
3	LB125789BSD	LB125789BSD	WATER	1.3	1000	100	3.2901	3.2901	3.04	3.3072	3.3072	0.0171	17.1
4	O2993-01	GRAB	WATER	1.6	1000	100	3.0387	3.0387	3.05	3.0681	3.0681	0.0294	29.4
5	O3019-01	MH-289	WATER	1.6	1000	100	3.0053	3.0053	3.04	3.0066	3.0066	0.0013	1.3

QC Batch# LB125789Test: TPHAnalysis Date: 06/05/2023**Chemicals Used:**

Chemical Name	Chemical Lot #
HEXANE	W3010
pH Paper 0-14	M4909
Sodium Sulfate	EP2340
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator DataAnalytical Balance ID # : WC SC-6Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 15:26
1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 14:30
Bal Check Time: 13:31 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 16:00
Out Time1: 15:25

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 17:01
1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 16:30
Bal Check Time: 17:40 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 17:35
Out Time2: 17:00

WORKLIST(Hardcopy Internal Chain)

WB 125789

WorkList Name : tph o3039

WorkList ID : 170554

Department : Wet-Chemistry

Date : 06-05-2023 12:13:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O2993-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	O11	05/31/2023	1664A
O3019-01	MH-289	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	D11	06/02/2023	1664A

Date/Time 06/05/23 12:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/05/23 15:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]