



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125792

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP102782
Turbidity Calibration std, 20NTU	WP102778
Turbidity Calibration std, 40NTU	WP102777
Turbidity Calibration std, 1NTU	WP102781
Turbidity Calibration std, 5NTU	WP102780
Turbidity Calibration - CCV std, 10 NTU	WP102779
10 NTU Standard 500 ml	W2971

Intercept: 0.1264

Slope: 0.9372

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.02		06/05/2023	09:15
2	CAL2	1	1	1.160	1.10	10.3	06/05/2023	09:19
3	CAL3	5	1	4.560	4.73	-5.4	06/05/2023	09:23
4	CAL4	10	1	9.620	10.13	1.3	06/05/2023	09:26
5	CAL5	20	1	18.900	20.03	0.2	06/05/2023	09:30
6	CAL6	40	1	37.600	39.99	-0	06/05/2023	09:34



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB125792

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.390	9.884	06/05/2023	09:38
2	ICB		1	0.146	0.021	06/05/2023	09:42
3	CCV1	10	1	9.280	9.767	06/05/2023	09:46
4	CCB1		1	0.148	0.023	06/05/2023	09:50
5	LB125792BL		1	0.152	0.027	06/05/2023	09:54
6	O3037-01		1	0.817	0.737	06/05/2023	10:00
7	O3037-01DUP		1	0.818	0.738	06/05/2023	10:04
8	O3037-03		1	0.207	0.086	06/05/2023	10:08
9	O3037-05		1	0.224	0.104	06/05/2023	10:12
10	O3037-07		5	17.800	18.858	06/05/2023	10:16
11	CCV2	10	1	9.360	9.852	06/05/2023	10:20
12	CCB2		1	0.146	0.021	06/05/2023	10:24

WORKLIST(Hardcopy Internal Chain)

LB125792

WorkList Name : TURB-060223

WorkList ID : 170519

Department : Wet-Chemistry

Date : 06-02-2023 14:44:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3037-01	MW-1	Water	Turbidity	Cool 4 deg C	LOCK01	D11	06/01/2023	SM2130 B
O3037-03	MW-3	Water	Turbidity	Cool 4 deg C	LOCK01	D11	06/01/2023	SM2130 B
O3037-05	MW-4	Water	Turbidity	Cool 4 deg C	LOCK01	D11	06/01/2023	SM2130 B
O3037-07	MW-2	Water	Turbidity	Cool 4 deg C	LOCK01	D11	06/01/2023	SM2130 B

Date/Time 06/05/23 08:50
 Raw Sample Received by: 12 (WC)
 Raw Sample Relinquished by: 18 (WC)

Date/Time 06/05/23 11:05
 Raw Sample Received by: 18 (WC)
 Raw Sample Relinquished by: 12 (WC)