

LB125794


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49985



A= 0.0000e+000

B= 2.0211e-004

C= 2.0169e-002

Rho= 0.9995428

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.008	0.008	-61	0.000	-61					
0.2	0.200	0.237	0.037	1075	0.0 %	1075					19
2.5	2.500	2.466	-0.034	12100	0.0 %	12100					-1
5.0	5.000	5.088	0.088	25076	0.0 %	25076					-2
7.5	7.500	7.278	-0.222	35911	0.0 %	35911					-3
10.0	10.000	10.123	0.123	49985	0.0 %	49985					1

LB125794 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jun 2023 15:29:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-61	-	0.0000	7471B	PPB	05 Jun 2023 16:11:28	S	Std
0.2 - 1	50.2	1075	-	0.2000	7471B	PPB	05 Jun 2023 16:13:45	S	Std
2.5 - 1	52.5	12100	-	2.5000	7471B	PPB	05 Jun 2023 16:16:02	S	Std
5.0 - 1	55	25076	-	5.0000	7471B	PPB	05 Jun 2023 16:18:20	S	Std
7.5 - 1	57.5	35911	-	7.5000	7471B	PPB	05 Jun 2023 16:23:15	S	Std
10.0 - 1	514	49985	-	10.0000	7471B	PPB	05 Jun 2023 16:25:32	S	Std
ICV06 - 1	ICV06	20190	4.1008	-	7471B	PPB	05 Jun 2023 16:30:54	U	SMPL
ICB06 - 1	ICB06	-153	-0.0108	-	7471B	PPB	05 Jun 2023 16:33:09	U	SMPL
CCV75 - 1	CCV75	25461	5.1661	-	7471B	PPB	05 Jun 2023 16:35:26	U	SMPL
CCB75 - 1	CCB75	-171	-0.0144	-	7471B	PPB	05 Jun 2023 16:37:42	U	SMPL
CRA - 1	CRA	1046	0.2316	-	7471B	PPB	05 Jun 2023 16:39:59	U	SMPL
HighStd - 1	HighStd	50435	10.2136	-	7471B	PPB	05 Jun 2023 16:42:14	U	SMPL
ChkStd - 1	ChkStd	35261	7.1468	-	7471B	PPB	05 Jun 2023 16:44:30	U	SMPL
PB153211BL - 1	PBS	-221	-0.0245	-	7471B	PPB	05 Jun 2023 16:46:49	U	SMPL
PB153211BS - 1	LCSS	22209	4.5088	-	7471B	PPB	05 Jun 2023 16:49:07	U	SMPL
O3018-03 - 1	362	94078	19.0342	-	7471B	PPB	05 Jun 2023 16:51:25	U	SMPL
O3018-05 - 1	348	100946	20.4223	-	7471B	PPB	05 Jun 2023 16:53:43	U	SMPL
O3018-07 - 1	343	98953	20.0195	-	7471B	PPB	05 Jun 2023 16:56:04	U	SMPL
O3018-09 - 1	72-11991	79665	16.1212	-	7471B	PPB	05 Jun 2023 16:58:28	U	SMPL
O3018-11 - 1	278	69059	13.9777	-	7471B	PPB	05 Jun 2023 17:00:48	U	SMPL
CCV76 - 1	CCV76	24957	5.0642	-	7471B	PPB	05 Jun 2023 17:03:06	U	SMPL
CCB76 - 1	CCB76	-138	-0.0077	-	7471B	PPB	05 Jun 2023 17:05:25	U	SMPL
O3018-13 - 1	RT-4648	139557	28.2260	-	7471B	PPB	05 Jun 2023 17:07:42	U	SMPL
O3026-01 - 1	VNJ-223	4977	1.0261	-	7471B	PPB	05 Jun 2023 17:09:58	U	SMPL
O3028-01 - 1	OR-02-060223	3783	0.7847	-	7471B	PPB	05 Jun 2023 17:12:24	U	SMPL
O3029-01 - 1	TP-6	496	0.1204	-	7471B	PPB	05 Jun 2023 17:14:41	U	SMPL
O3029-04 - 1	TP-7	652	0.1519	-	7471B	PPB	05 Jun 2023 17:16:56	U	SMPL
O3029-07 - 1	TP-8	496	0.1204	-	7471B	PPB	05 Jun 2023 17:19:13	U	SMPL
O3029-10 - 1	TP-9	618	0.1451	-	7471B	PPB	05 Jun 2023 17:21:30	U	SMPL
O3029-13 - 1	TP-10	665	0.1546	-	7471B	PPB	05 Jun 2023 17:23:47	U	SMPL
O3032-01 - 1	TP-20-0A	471	0.1154	-	7471B	PPB	05 Jun 2023 17:26:04	U	SMPL
O3032-04 - 1	TP-21-0A	334	0.0877	-	7471B	PPB	05 Jun 2023 17:28:20	U	SMPL
CCV77 - 1	CCV77	25188	5.1109	-	7471B	PPB	05 Jun 2023 17:30:36	U	SMPL
CCB77 - 1	CCB77	-93	0.0014	-	7471B	PPB	05 Jun 2023 17:32:52	U	SMPL
O3036-01 - 1	PL-STONE-0602	19	0.0240	-	7471B	PPB	05 Jun 2023 17:35:09	U	SMPL
O3038-01 - 1	NB-07-060223	1709	0.3656	-	7471B	PPB	05 Jun 2023 17:37:25	U	SMPL
O3038-01DUP - 1	NB-07-060223DUP	1657	0.3551	-	7471B	PPB	05 Jun 2023 17:39:41	U	SMPL
O3038-01MS - 1	NB-07-060223MS	22581	4.5840	-	7471B	PPB	05 Jun 2023 17:44:31	U	SMPL
O3038-01MSD - 1	NB-07-060223MSD	20672	4.1982	-	7471B	PPB	05 Jun 2023 17:46:47	U	SMPL
O3038-01LX5 - 1		211	0.0628	-	7471B	PPB	05 Jun 2023 17:49:05	U	SMPL
O3038-01A - 1		22824	4.6331	-	7471B	PPB	05 Jun 2023 17:51:23	U	SMPL
CCV78 - 1	CCV78	24899	5.0525	-	7471B	PPB	05 Jun 2023 17:53:40	U	SMPL
CCB78 - 1	CCB78	-122	-0.0045	-	7471B	PPB	05 Jun 2023 17:55:59	U	SMPL
O3018-03DLX5 - 1		16521	3.3592	-	7471B	PPB	05 Jun 2023 17:58:16	U	SMPL
O3018-05DLX5 - 1		20338	4.1307	-	7471B	PPB	05 Jun 2023 18:00:32	U	SMPL
O3018-07DLX5 - 1		19816	4.0252	-	7471B	PPB	05 Jun 2023 18:02:49	U	SMPL
O3018-09DLX5 - 1		15378	3.1282	-	7471B	PPB	05 Jun 2023 18:05:06	U	SMPL
O3018-11DLX5 - 1		13775	2.8042	-	7471B	PPB	05 Jun 2023 18:07:24	U	SMPL
O3018-13DLX5 - 1		27104	5.4982	-	7471B	PPB	05 Jun 2023 18:09:41	U	SMPL
CCV79 - 1	CCV79	25786	5.2318	-	7471B	PPB	05 Jun 2023 18:11:58	U	SMPL
CCB79 - 1	CCB79	-139	-0.0079	-	7471B	PPB	05 Jun 2023 18:14:16	U	SMPL