



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125798

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	06/05/2023	15:10
2	O3023-01		1	12	100.00	>212.0	>212.0	06/05/2023	15:40
3	O3023-01DUP		1	12	100.00	>212.0	>212.0	06/05/2023	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 125798

WorkList Name : fp s 6-5

WorkList ID : 170537

Department : Wet-Chemistry

Date : 06-05-2023 08:29:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3023-01 A	32415-32417	Solid	Flash Point	Cool 4 deg C	PSEG03	D11	06/02/2023	1010B

Date/Time 06/05/2023 15:00
 Raw Sample Received by: Ruy Casp
 Raw Sample Relinquished by: JDC

Date/Time 06/05/2023 16:45
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: Ruy Casp