



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125805

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/06/2023	09:45
2	CAL2	1	Water	NA	NA	20.7	7.00	06/06/2023	09:46
3	CAL3	1	Water	NA	NA	20.7	10.02	06/06/2023	09:47
4	ICV	1	Water	NA	NA	20.8	6.99	06/06/2023	09:50
5	CCV1	1	Water	NA	NA	20.6	2.01	06/06/2023	10:00
6	O3046-01	1	Water	NA	NA	21.7	7.04	06/06/2023	10:01
7	O3046-01DUP	1	Water	NA	NA	21.8	7.05	06/06/2023	10:02
8	O3046-02	1	Water	NA	NA	21.7	7.89	06/06/2023	10:10
9	O3046-03	1	Water	NA	NA	22.0	7.10	06/06/2023	10:14
10	O3046-04	1	Water	NA	NA	22.0	6.67	06/06/2023	10:15
11	O3046-05	1	Water	NA	NA	22.1	6.33	06/06/2023	10:20
12	O3046-06	1	Water	NA	NA	21.9	6.97	06/06/2023	10:22
13	O3046-07	1	Water	NA	NA	22.3	6.92	06/06/2023	10:25
14	O3046-08	1	Water	NA	NA	22.2	6.91	06/06/2023	10:30
15	O3046-09	1	Water	NA	NA	22.1	6.76	06/06/2023	10:33
16	CCV2	1	Water	NA	NA	20.8	12.02	06/06/2023	10:35
17	O3046-10	1	Water	NA	NA	22.2	7.13	06/06/2023	10:41
18	O3046-11	1	Water	NA	NA	22.3	6.93	06/06/2023	10:45
19	O3046-12	1	Water	NA	NA	22.3	6.51	06/06/2023	10:48
20	O3046-13	1	Water	NA	NA	22.4	4.12	06/06/2023	10:55

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O3046-14	1	Water	NA	NA	22.3	3.45	06/06/2023	11:00
22	O3046-15	1	Water	NA	NA	22.4	6.76	06/06/2023	11:10
23	O3046-16	1	Water	NA	NA	22.3	6.36	06/06/2023	11:15
24	O3046-17	1	Water	NA	NA	22.2	7.78	06/06/2023	11:20
25	O3046-18	1	Water	NA	NA	22.5	8.36	06/06/2023	11:25
26	CCV3	1	Water	NA	NA	20.8	2.02	06/06/2023	11:30

✓ B125805

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 03047

WorkList ID : 170582

Department : Wet-Chemistry

Date : 06-06-2023 09:32:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3046-01	COMP-1-1-5	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-02	1	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-03	2	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-04	3	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-05	4	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-06	5	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-07	COMP-2-6-10	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-08	6	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-09	7	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-10	8	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-11	9	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-12	10	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-13	COMP-3-11-15	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-14	11	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-15	12	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-16	13	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-17	14	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C
O3046-18	15	Water	pH	Cool 4 deg C	PSEG03	D11	06/05/2023	9040C

06/06/23 09:40

Date/Time

Raw Sample Received by:

W1W6J

Raw Sample Relinquished by:

W1W6J

Date/Time 06/06/23

Raw Sample Received by:

W1W6J

Raw Sample Relinquished by:

W1W6J