

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125806

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/06/2023	09:45
2	CAL2	1	Water	NA	NA	20.7	7.00	06/06/2023	09:46
3	CAL3	1	Water	NA	NA	20.7	10.02	06/06/2023	09:47
4	ICV	1	Water	NA	NA	20.8	6.99	06/06/2023	09:50
5	CCV1	1	Water	NA	NA	20.6	2.02	06/06/2023	11:35
6	O3039-06	1	Solid	20.04	20	25.5	5.72	06/06/2023	11:40
7	O3044-04	1	Solid	20.02	20	22.0	9.08	06/06/2023	11:45
8	O3044-07	1	Solid	20.01	20	20.9	9.25	06/06/2023	11:48
9	O3044-10	1	Solid	20.04	20	21.5	8.51	06/06/2023	11:50
10	O3044-13	1	Solid	20.02	20	21.2	8.81	06/06/2023	12:00
11	O3044-16	1	Solid	20.03	20	21.2	8.59	06/06/2023	12:05
12	O3047-03	1	Solid	20.02	20	21.2	8.59	06/06/2023	12:10
13	O3047-06	1	Solid	20.03	20	21.5	8.49	06/06/2023	12:15
14	O3050-02	1	Solid	20.04	20	22.7	6.42	06/06/2023	12:20
15	O3050-05	1	Solid	20.05	20	23.1	8.62	06/06/2023	12:25
16	CCV2	1	Water	NA	NA	20.9	12.01	06/06/2023	12:30
17	O3050-05DUP	1	Solid	20.04	20	23.2	8.63	06/06/2023	12:31
18	CCV3	1	Water	NA	NA	20.7	2.01	06/06/2023	12:35

VB 125806

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRISIVITY O3057 WorkList ID : 170585 Department : Wet-Chemistry Date : 06-06-2023 09:33:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3039-06	WC31	Solid	Corrosivity	Cool 4 deg C	loui01	D11	06/02/2023	9045D
O3044-04	SB-02-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D11	06/02/2023	9045D
O3044-07	SB-03-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D11	06/02/2023	9045D
O3044-10	SB-04-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D11	06/02/2023	9045D
O3044-13	SB-07-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D11	06/02/2023	9045D
O3044-16	SB-08-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D11	06/02/2023	9045D
O3047-03	WC32	Solid	Corrosivity	Cool 4 deg C	loui01	D11	06/02/2023	9045D
O3047-06	WC33	Solid	Corrosivity	Cool 4 deg C	loui01	D11	06/05/2023	9045D
O3050-02	TP-22-0A	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	06/05/2023	9045D
O3050-05	TP-23-0A	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	06/05/2023	9045D

Date/Time 06/06/23 09:35
 Raw Sample Received by: 106091
 Raw Sample Relinquished by: [Signature]

Date/Time 06/06/23 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]