

LB125808


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

59182



A= 0.0000e+000

B= 1.6807e-004

C= -4.6702e-003

Rho= 0.9997692

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	4/1/11
0.0	0.000	0.049	0.049	322	0.000	322					
0.2	0.200	0.215	0.015	1307	0.0 %	1307					
2.5	2.500	2.353	-0.147	14027	0.0 %	14027					
5.0	5.000	5.059	0.059	30130	0.0 %	30130					
7.5	7.500	7.581	0.081	45132	0.0 %	45132					
10.0	10.000	9.942	-0.058	59182	0.0 %	59182					

4/1/11
 7
 8 -6
 1 MB
 1 618123
 -1

LB125808 INSTRUMENT ID :CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Jun 2023 12:49:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	322	-	0.0000	7471B	PPB	06 Jun 2023 13:38:33	S	Std
0.2 - 1	50.2	1307	-	0.2000	7471B	PPB	06 Jun 2023 13:40:54	S	Std
2.5 - 1	52.5	14027	-	2.5000	7471B	PPB	06 Jun 2023 13:43:12	S	Std
5.0 - 1	55	30130	-	5.0000	7471B	PPB	06 Jun 2023 13:45:29	S	Std
7.5 - 1	57.5	45132	-	7.5000	7471B	PPB	06 Jun 2023 13:47:46	S	Std
10.0 - 1	510	59182	-	10.0000	7471B	PPB	06 Jun 2023 13:57:55	S	Std
ICV07 - 1	ICV07	23924	4.0164	-	7471B	PPB	06 Jun 2023 14:00:56	U	SMPL
ICB07 - 1	ICB07	-43	-0.0119	-	7471B	PPB	06 Jun 2023 14:03:11	U	SMPL
CCV80 - 1	CCV80	30039	5.0441	-	7471B	PPB	06 Jun 2023 14:05:28	U	SMPL
CCB80 - 1	CCB80	-83	-0.0186	-	7471B	PPB	06 Jun 2023 14:07:44	U	SMPL
CRA - 1	CRA	1195	0.1962	-	7471B	PPB	06 Jun 2023 14:10:01	U	SMPL
HighStd - 1	HighStd	58541	9.8346	-	7471B	PPB	06 Jun 2023 14:12:16	U	SMPL
ChkStd - 1	ChkStd	41939	7.0442	-	7471B	PPB	06 Jun 2023 14:14:31	U	SMPL
PB153243BL - 1	PBS	-167	-0.0327	-	7471B	PPB	06 Jun 2023 14:16:50	U	SMPL
PB153243BS - 1	LCSS	24438	4.1027	-	7471B	PPB	06 Jun 2023 14:19:09	U	SMPL
O3039-01 - 1	SB30A	9917	1.6621	-	7471B	PPB	06 Jun 2023 14:21:26	U	SMPL
O3039-02 - 1	SB30B	2030	0.3365	-	7471B	PPB	06 Jun 2023 14:23:44	U	SMPL
O3039-04 - 1	SB31A	12149	2.0373	-	7471B	PPB	06 Jun 2023 14:26:02	U	SMPL
O3039-05 - 1	SB31B	323	0.0496	-	7471B	PPB	06 Jun 2023 14:28:20	U	SMPL
O3045-01 - 1	PAT-1-060523	1266	0.2081	-	7471B	PPB	06 Jun 2023 14:30:36	U	SMPL
CCV81 - 1	CCV81	28754	4.8282	-	7471B	PPB	06 Jun 2023 14:32:52	U	SMPL
CCB81 - 1	CCB81	-21	-0.0082	-	7471B	PPB	06 Jun 2023 14:35:07	U	SMPL
O3050-01 - 1	TP-22-0A	456	0.0720	-	7471B	PPB	06 Jun 2023 14:37:24	U	SMPL
O3050-04 - 1	TP-23-0A	600	0.0962	-	7471B	PPB	06 Jun 2023 14:39:40	U	SMPL
O3050-04DUP - 1	TP-23-0ADUP	633	0.1017	-	7471B	PPB	06 Jun 2023 14:41:56	U	SMPL
O3050-04MS - 1	TP-23-0AMS	24680	4.1434	-	7471B	PPB	06 Jun 2023 14:44:12	U	SMPL
O3050-04MSD - 1	TP-23-0AMSD	24732	4.1522	-	7471B	PPB	06 Jun 2023 14:46:28	U	SMPL
O3050-04LX5 - 1		-56	-0.0141	-	7471B	PPB	06 Jun 2023 14:48:46	U	SMPL
O3050-04A - 1		23534	3.9508	-	7471B	PPB	06 Jun 2023 14:51:06	U	SMPL
CCV82 - 1	CCV82	28361	4.7621	-	7471B	PPB	06 Jun 2023 14:53:23	U	SMPL
CCB82 - 1	CCB82	-106	-0.0225	-	7471B	PPB	06 Jun 2023 14:55:42	U	SMPL