

LB125813


245.1

INSTRUMENT NO: CV1

Linear

 μ Abs.:

39362



A= 0.0000e+000

B= 2.5339e-004

C= 5.1766e-004

Rho= 0.9999301

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	0.019	0.019	73	0.000	73					1
0.05	0.050	0.056	0.006	220	0.0 %	220	0				12
0.2	0.200	0.212	0.012	835	0.0 %	835					6
2.5	2.500	2.465	-0.035	9728	0.0 %	9728					1
5.0	5.000	4.938	-0.062	19485	0.0 %	19485					1
7.5	7.500	7.585	0.085	29932	0.0 %	29932					1
10.0	10.000	9.974	-0.026	39362	0.0 %	39362					0

LB125813 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 06 Jun 2023 15:32:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	73	-	0.0000	245.1	PPB	06 Jun 2023 16:05:04	S	Std
0.05 - 1	50.05	220	-	0.0500	245.1	PPB	06 Jun 2023 16:07:22	S	Std
0.2 - 1	50.2	835	-	0.2000	245.1	PPB	06 Jun 2023 16:09:39	S	Std
2.5 - 1	52.5	9728	-	2.5000	245.1	PPB	06 Jun 2023 16:11:56	S	Std
5.0 - 1	53	19485	-	5.0000	245.1	PPB	06 Jun 2023 16:14:13	S	Std
7.5 - 1	57.5	29932	-	7.5000	245.1	PPB	06 Jun 2023 16:16:30	S	Std
10.0 - 1	510	39362	-	10.0000	245.1	PPB	06 Jun 2023 16:29:18	S	Std
ICV08 - 1	ICV08	15487	3.9247	-245.1		PPB	06 Jun 2023 16:43:53	U	SMPL
ICB08 - 1	ICB08	-154	-0.0385	-245.1		PPB	06 Jun 2023 16:46:08	U	SMPL
CCV83 - 1	CCV83	19729	4.9996	-245.1		PPB	06 Jun 2023 16:51:01	U	SMPL
CCB83 - 1	CCB83	-167	-0.0418	-245.1		PPB	06 Jun 2023 16:53:16	U	SMPL
CRA - 1	CRA	162	0.0416	-245.1		PPB	06 Jun 2023 16:58:10	U	SMPL
HighStd - 1	HighStd	40120	10.1665	-245.1		PPB	06 Jun 2023 17:03:06	U	SMPL
ChkStd - 1	ChkStd	28738	7.2824	-245.1		PPB	06 Jun 2023 17:08:07	U	SMPL
PB153245BL - 1	PBW	-206	-0.0517	-245.1		PPB	06 Jun 2023 17:10:23	U	SMPL
PB153245BS - 1	LCSW	16295	4.1295	-245.1		PPB	06 Jun 2023 17:12:41	U	SMPL
O3016-08 - 1	B-45(OW)	351	0.0895	-245.1		PPB	06 Jun 2023 17:14:57	U	SMPL
O3016-08DUP - 1	B-45(OW)DUP	308	0.0786	-245.1		PPB	06 Jun 2023 17:17:14	U	SMPL
O3016-08MS - 1	B-45(OW)MS	17516	4.4389	-245.1		PPB	06 Jun 2023 17:19:32	U	SMPL
O3016-08MSD - 1	B-45(OW)MSD	16377	4.1503	-245.1		PPB	06 Jun 2023 17:21:49	U	SMPL
O3039-09 - 1	MW-34	-32	-0.0076	-245.1		PPB	06 Jun 2023 17:24:06	U	SMPL
CCV84 - 1	CCV84	20692	5.2436	-245.1		PPB	06 Jun 2023 17:26:23	U	SMPL
CCB84 - 1	CCB84	-137	-0.0342	-245.1		PPB	06 Jun 2023 17:28:38	U	SMPL
O3047-09 - 1	MW-41	127	0.0327	-245.1		PPB	06 Jun 2023 17:30:55	U	SMPL
O3047-12 - 1	TWP-07	55	0.0145	-245.1		PPB	06 Jun 2023 17:33:11	U	SMPL
O3016-08LX5 - 1		-31	-0.0073	-245.1		PPB	06 Jun 2023 17:35:27	U	SMPL
O3016-08A - 1		16565	4.1979	-245.1		PPB	06 Jun 2023 17:40:14	U	SMPL
CCV85 - 1	CCV85	19087	4.8369	-245.1		PPB	06 Jun 2023 17:42:30	U	SMPL
CCB85 - 1	CCB85	-132	-0.0329	-245.1		PPB	06 Jun 2023 17:44:48	U	SMPL