



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125817

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	06/06/2023	13:00
2	O3016-08		1	12	100.00	>212.0	>212.0	06/06/2023	13:30
3	O3016-08DUP		1	12	100.00	>212.0	>212.0	06/06/2023	14:00
4	O3019-01		1	14	100.00	>212.0	>212.0	06/06/2023	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6125817

WorkList Name : FP 6-5

WorkList ID : 170544

Department : Wet-Chemistry

Date : 06-05-2023 08:49:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3016-08 <i>M</i>	B-45(OW)	Water	Flash Point	Cool 4 deg C	Iou01	D11	06/01/2023	1010B
O3019-01 <i>M</i>	MH-289	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/02/2023	1010B

Date/Time 06/06/2023 12:45
 Raw Sample Received by: RM awg
 Raw Sample Relinquished by: 78(awg)

Date/Time 06/06/2023 15:10
 Raw Sample Received by: 78(awg)
 Raw Sample Relinquished by: RM awg