

LB125836


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57687



A= 0.0000e+000

B= 1.6410e-004

C= -3.2092e-002

Rho= 0.9955516

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/L
0.0	0.000	-0.031	-0.031	5	0.000	5					
0.2	0.200	0.174	-0.026	1255	0.0 %	1255					13
2.5	2.500	2.306	-0.194	14248	0.0 %	14248					8
5.0	5.000	5.283	0.283	32387	0.0 %	32387					6
7.5	7.500	8.035	0.535	49159	0.0 %	49159					7
10.0	10.000	9.434	-0.566	57687	0.0 %	57687					6

LB125836 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Jun 2023 14:47:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	7471B	PPB	07 Jun 2023 14:49:13	S	Std
0.2 - 1	50.2	1255	-	0.2000	7471B	PPB	07 Jun 2023 14:51:31	S	Std
2.5 - 1	52.5	14248	-	2.5000	7471B	PPB	07 Jun 2023 14:53:48	S	Std
5.0 - 1	55	32387	-	5.0000	7471B	PPB	07 Jun 2023 14:56:05	S	Std
7.5 - 1	57.5	49159	-	7.5000	7471B	PPB	07 Jun 2023 14:58:23	S	Std
10.0 - 1	511	57687	-	10.0000	7471B	PPB	07 Jun 2023 15:00:40	S	Std
ICV10 - 1	ICV10	24153	3.9313	-	7471B	PPB	07 Jun 2023 15:03:30	U	SMPL
ICB10 - 1	ICB10	-158	-0.0580	-	7471B	PPB	07 Jun 2023 15:05:46	U	SMPL
CCV90 - 1	CCV90	30408	4.9578	-	7471B	PPB	07 Jun 2023 15:08:03	U	SMPL
CCB90 - 1	CCB90	-183	-0.0621	-	7471B	PPB	07 Jun 2023 15:10:18	U	SMPL
CRA - 1	CRA	1221	0.1683	-	7471B	PPB	07 Jun 2023 15:12:35	U	SMPL
HighStd - 1	HighStd	57238	9.3605	-	7471B	PPB	07 Jun 2023 15:14:51	U	SMPL
ChkStd - 1	ChkStd	43964	7.1823	-	7471B	PPB	07 Jun 2023 15:17:06	U	SMPL
PB153288BL - 1	PB153288BL	-317	-0.0841	-	7471B	PPB	07 Jun 2023 15:19:25	U	SMPL
PB153288BS - 1	PB153288BS	21671	3.5241	-	7471B	PPB	07 Jun 2023 15:21:44	U	SMPL
O3042-01 - 1	WC-1	16089	2.6081	-	7471B	PPB	07 Jun 2023 15:24:01	U	SMPL
O3047-01 - 1	SB32A	5827	0.9241	-	7471B	PPB	07 Jun 2023 15:26:19	U	SMPL
O3047-02 - 1	SB32B	3258	0.5025	-	7471B	PPB	07 Jun 2023 15:28:38	U	SMPL
O3047-04 - 1	SB33A	11044	1.7802	-	7471B	PPB	07 Jun 2023 15:30:56	U	SMPL
O3047-05 - 1	SB33B	320	0.0204	-	7471B	PPB	07 Jun 2023 15:33:12	U	SMPL
CCV91 - 1	CCV91	30354	4.9489	-	7471B	PPB	07 Jun 2023 15:35:29	U	SMPL
CCB91 - 1	CCB91	-91	-0.0470	-	7471B	PPB	07 Jun 2023 15:37:44	U	SMPL
O3049-01 - 1	HR-01-060523	8471	1.3580	-	7471B	PPB	07 Jun 2023 15:40:01	U	SMPL
O3049-01MS - 1	HR-01-060523MS	28592	4.6598	-	7471B	PPB	07 Jun 2023 15:44:33	U	SMPL
O3049-01MSD - 1	HR-01-060523MSD	31232	5.0930	-	7471B	PPB	07 Jun 2023 15:46:49	U	SMPL
O3049-03 - 1	HR-03-060523	5435	0.8598	-	7471B	PPB	07 Jun 2023 15:49:07	U	SMPL
O3051-02 - 1	RBR-251580	4080	0.6374	-	7471B	PPB	07 Jun 2023 15:51:26	U	SMPL
O3064-01 - 1	CHRT-24236	12261	1.9799	-	7471B	PPB	07 Jun 2023 15:53:42	U	SMPL
O3068-01 - 1	SU-2-060623	3498	0.5419	-	7471B	PPB	07 Jun 2023 15:55:59	U	SMPL
O3049-01DUP - 1	HR-01-060523DUP	8324	1.3339	-	7471B	PPB	07 Jun 2023 15:59:16	U	SMPL
PB153289BL - 1	PBS	0	-0.0321	-	7471B	PPB	07 Jun 2023 16:01:48	U	SMPL
PB153289BS - 1	LCSS	23488	3.8222	-	7471B	PPB	07 Jun 2023 16:04:03	U	SMPL
CCV92 - 1	CCV92	30541	4.9796	-	7471B	PPB	07 Jun 2023 16:06:19	U	SMPL
CCB92 - 1	CCB92	-126	-0.0528	-	7471B	PPB	07 Jun 2023 16:08:36	U	SMPL
O3044-01 - 1	SB-02-0-2.0	25760	4.1950	-	7471B	PPB	07 Jun 2023 16:10:53	U	SMPL
O3044-02 - 1	SB-02-5-7.0	32733	5.3393	-	7471B	PPB	07 Jun 2023 16:13:10	U	SMPL
O3044-03 - 1	SB-02-5.0-7.0-DUP	26173	4.2628	-	7471B	PPB	07 Jun 2023 16:15:27	U	SMPL
O3044-05 - 1	SB-03-0-2.0	52408	8.5679	-	7471B	PPB	07 Jun 2023 16:17:45	U	SMPL
O3044-06 - 1	SB-03-5.0-7.0	187701	30.7691	-	7471B	PPB	07 Jun 2023 16:20:02	U	SMPL
O3044-08 - 1	SB-04-0-2.0	33693	5.4968	-	7471B	PPB	07 Jun 2023 16:22:21	U	SMPL
O3044-09 - 1	SB-04-5.0-7.0	59220	9.6857	-	7471B	PPB	07 Jun 2023 16:24:47	U	SMPL
O3044-11 - 1	SB-07-0-2.0	34203	5.5805	-	7471B	PPB	07 Jun 2023 16:27:06	U	SMPL
O3044-12 - 1	SB-07-5.0-7.0	33165	5.4102	-	7471B	PPB	07 Jun 2023 16:29:26	U	SMPL
O3044-14 - 1	SB-08-0-2.0	34387	5.6107	-	7471B	PPB	07 Jun 2023 16:31:44	U	SMPL
CCV93 - 1	CCV93	30523	4.9766	-	7471B	PPB	07 Jun 2023 16:34:01	U	SMPL
CCB93 - 1	CCB93	-163	-0.0588	-	7471B	PPB	07 Jun 2023 16:36:19	U	SMPL
O3044-15 - 1	SB-08-5.0-7.0	22792	3.7080	-	7471B	PPB	07 Jun 2023 16:38:36	U	SMPL
O3044-15DUP - 1	SB-08-5.0-7.0DUP	20695	3.3639	-	7471B	PPB	07 Jun 2023 16:40:53	U	SMPL
O3044-15MS - 1	SB-08-5.0-7.0MS	42219	6.8959	-	7471B	PPB	07 Jun 2023 16:43:10	U	SMPL
O3044-15MSD - 1	SB-08-5.0-7.0MSD	43380	7.0864	-	7471B	PPB	07 Jun 2023 16:45:27	U	SMPL
O3049-01LX5 - 1		1411	0.1994	-	7471B	PPB	07 Jun 2023 16:47:46	U	SMPL
O3049-01A - 1		30137	4.9133	-	7471B	PPB	07 Jun 2023 16:50:05	U	SMPL
O3044-15LX5 - 1		4403	0.6904	-	7471B	PPB	07 Jun 2023 16:52:21	U	SMPL
O3044-15A - 1		43246	7.0645	-	7471B	PPB	07 Jun 2023 16:54:40	U	SMPL
O3044-06DLX5 - 1		37885	6.1847	-	7471B	PPB	07 Jun 2023 16:56:57	U	SMPL
CCV94 - 1	CCV94	31582	5.1504	-	7471B	PPB	07 Jun 2023 16:59:15	U	SMPL
CCB94 - 1	CCB94	-159	-0.0582	-	7471B	PPB	07 Jun 2023 17:01:32	U	SMPL