



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125839

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	06/07/2023	15:30
2	CAL2	1	Water	NA	NA	20.9	7.00	06/07/2023	15:31
3	CAL3	1	Water	NA	NA	20.8	10.02	06/07/2023	15:33
4	ICV	1	Water	NA	NA	20.6	7.00	06/07/2023	15:35
5	CCV1	1	Water	NA	NA	20.7	2.01	06/07/2023	15:40
6	O3069-03	1	Solid	20.02	20	20.2	7.82	06/07/2023	15:44
7	O3069-06	1	Solid	20.01	20	20.3	7.31	06/07/2023	15:47
8	O3069-08	1	Solid	20.03	20	20.7	7.52	06/07/2023	15:50
9	O3092-02	1	Solid	20.02	20	22.6	7.62	06/07/2023	16:00
10	O3092-05	1	Solid	20.03	20	21.9	7.49	06/07/2023	16:05
11	O3092-05DUP	1	Solid	20.04	20	21.8	7.50	06/07/2023	16:06
12	CCV2	1	Water	NA	NA	20.8	12.02	06/07/2023	16:10

WORKLIST(Hardcopy Internal Chain)

VB 125839

WorkList Name : CORRRSIVITY O3092

WorkList ID : 170634

Department : Wet-Chemistry

Date : 06-07-2023 13:37:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3069-03	WC34	Solid	Corrosivity	Cool 4 deg C	loui01	D11	06/06/2023	9045D
O3069-06	WC35	Solid	Corrosivity	Cool 4 deg C	loui01	D11	06/06/2023	9045D
O3069-08	WC36	Solid	Corrosivity	Cool 4 deg C	loui01	D11	06/06/2023	9045D
O3092-02	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	06/07/2023	9045D
O3092-05	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	06/07/2023	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

06/07/23 15:00

JPWCC

JPWCC

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

06/07/23

JPWCC

JPWCC

17:30

JPWCC

JPWCC