



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125841

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/07/2023	15:25
2	CAL2	1	Water	NA	NA	20.8	6.99	06/07/2023	15:26
3	CAL3	1	Water	NA	NA	20.8	10.03	06/07/2023	15:30
4	ICV	1	Water	NA	NA	20.7	6.98	06/07/2023	15:33
5	CCV1	1	Water	NA	NA	20.6	2.01	06/07/2023	15:37
6	O3066-01	1	Water	NA	NA	22.0	6.98	06/07/2023	15:40
7	O3066-01DUP	1	Water	NA	NA	22.1	6.99	06/07/2023	15:42
8	O3091-01	1	Water	NA	NA	20.5	7.11	06/07/2023	15:45
9	CCV2	1	Water	NA	NA	20.8	12.02	06/07/2023	15:50

VB 125841

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o3091

WorkList ID : 170641

Department : Wet-Chemistry

Date : 06-07-2023 15:00:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3066-01	C 606	Water	pH	Cool 4 deg C	PSEG03	D11	06/06/2023	9040C
O3091-01	B 402	Water	pH	Cool 4 deg C	PSEG04	D11	06/07/2023	9040C

Date/Time 06/07/23 15:16

Raw Sample Received by: JPC

Raw Sample Relinquished by: JPC

Date/Time 06/07/23

Raw Sample Received by: JPC

Raw Sample Relinquished by: JPC