

L3125842


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

43091



A= 0.0000e+000

B= 2.2798e-004

C= -1.9314e-002

Rho= 0.9993474

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/1)
0.0	0.000	-0.004	-0.004	66	0.000	66	0				-
0.2	0.200	0.193	-0.007	930	0.0 %	930					-4
2.5	2.500	2.393	-0.107	10581	0.0 %	10581					-4
5.0	5.000	5.095	0.095	22435	0.0 %	22435					2
7.5	7.500	7.719	0.219	33941	0.0 %	33941					3
10.0	10.000	9.805	-0.195	43091	0.0 %	43091					-2

LB125842 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jun 2023 17:18:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	66	-	0.0000	7470A	PPB	07 Jun 2023 17:18:43	S	Std
0.2 - 1	50.2	930	-	0.2000	7470A	PPB	07 Jun 2023 17:21:00	S	Std
2.5 - 1	52.5	10581	-	2.5000	7470A	PPB	07 Jun 2023 17:23:17	S	Std
5.0 - 1	55	22435	-	5.0000	7470A	PPB	07 Jun 2023 17:25:35	S	Std
7.5 - 1	57.5	33941	-	7.5000	7470A	PPB	07 Jun 2023 17:27:51	S	Std
10.0 - 1	514	43091	-	10.0000	7470A	PPB	07 Jun 2023 17:30:08	S	Std
ICV11 - 1	ICV11	16195	3.6728	-	7470A	PPB	07 Jun 2023 17:35:34	U	SMPL
ICB11 - 1	ICB11	-32	-0.0266	-	7470A	PPB	07 Jun 2023 17:37:49	U	SMPL
CCV95 - 1	CCV95	22278	5.0596	-	7470A	PPB	07 Jun 2023 17:40:05	U	SMPL
CCB95 - 1	CCB95	-55	-0.0319	-	7470A	PPB	07 Jun 2023 17:42:20	U	SMPL
CRA - 1	CRA	924	0.1913	-	7470A	PPB	07 Jun 2023 17:44:36	U	SMPL
HighStd - 1	HighStd	42473	9.6637	-	7470A	PPB	07 Jun 2023 17:46:52	U	SMPL
ChkStd - 1	ChkStd	30951	7.0369	-	7470A	PPB	07 Jun 2023 17:49:08	U	SMPL
PB153257BL - 1	PBW	-129	-0.0487	-	7470A	PPB	07 Jun 2023 17:51:26	U	SMPL
PB153257BS - 1	LCSW	18596	4.2202	-	7470A	PPB	07 Jun 2023 17:53:44	U	SMPL
O3024-01 - 1	TOPSOIL-OU3-GRILLO-1	771	0.1565	-	7470A	PPB	07 Jun 2023 17:56:00	U	SMPL
O3024-01DUP - 1	TOPSOIL-OU3-GRILLO-1DUP	817	0.1669	-	7470A	PPB	07 Jun 2023 17:58:18	U	SMPL
O3024-01MS - 1	TOPSOIL-OU3-GRILLO-1MS	19212	4.3607	-	7470A	PPB	07 Jun 2023 18:00:35	U	SMPL
O3024-01MSD - 1	TOPSOIL-OU3-GRILLO-1MSD	19618	4.4532	-	7470A	PPB	07 Jun 2023 18:02:53	U	SMPL
CCV96 - 1	CCV96	22318	5.0688	-	7470A	PPB	07 Jun 2023 18:05:09	U	SMPL
CCB96 - 1	CCB96	-82	-0.0380	-	7470A	PPB	07 Jun 2023 18:07:25	U	SMPL
O3024-01LX5 - 1		205	0.0274	-	7470A	PPB	07 Jun 2023 18:09:42	U	SMPL
O3024-01A - 1		18191	4.1279	-	7470A	PPB	07 Jun 2023 18:11:58	U	SMPL
CCV97 - 1	CCV97	22742	5.1654	-	7470A	PPB	07 Jun 2023 18:14:13	U	SMPL
CCB97 - 1	CCB97	-103	-0.0428	-	7470A	PPB	07 Jun 2023 18:16:29	U	SMPL