

LB125247


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4513E



A= 0.0000e+000

B= 2.1820e-004

C= -1.4994e-002

Rho= 0.9992887

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.0	0.000	0.004	0.004	85	0.000	85	0				1
0.2	0.200	0.207	0.007	1016	0.0 %	1016					3
2.5	2.500	2.324	-0.176	10720	0.0 %	10720					-7
5.0	5.000	5.202	0.202	23910	0.0 %	23910					4
7.5	7.500	7.630	0.130	35035	0.0 %	35035					2
10.0	10.000	9.834	-0.166	45138	0.0 %	45138					-2

LB125847 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jun 2023 11:06:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	85	-	0.0000	7470A	PPB	08 Jun 2023 11:10:03	S	Std
0.2 - 1	502	1016	-	0.2000	7470A	PPB	08 Jun 2023 11:12:21	S	Std
2.5 - 1	52.5	10720	-	2.5000	7470A	PPB	08 Jun 2023 11:14:39	S	Std
5.0 - 1	55	23910	-	5.0000	7470A	PPB	08 Jun 2023 11:16:56	S	Std
7.5 - 1	57.5	35035	-	7.5000	7470A	PPB	08 Jun 2023 11:21:45	S	Std
10.0 - 1	510	45138	-	10.0000	7470A	PPB	08 Jun 2023 11:26:50	S	Std
ICV12 - 1	ICV12	17425	3.7871	-	7470A	PPB	08 Jun 2023 11:29:40	U	SMPL
ICB12 - 1	ICB12	-203	-0.0606	-	7470A	PPB	08 Jun 2023 11:31:55	U	SMPL
CCV01 - 1	CCV01	24026	5.2274	-	7470A	PPB	08 Jun 2023 11:34:12	U	SMPL
CCB01 - 1	CCB01	-202	-0.0591	-	7470A	PPB	08 Jun 2023 11:36:27	U	SMPL
CRA - 1	CRA	894	0.1801	-	7470A	PPB	08 Jun 2023 11:38:44	U	SMPL
HighStd - 1	HighStd	44842	9.7694	-	7470A	PPB	08 Jun 2023 11:40:59	U	SMPL
ChkStd - 1	ChkStd	32439	7.0631	-	7470A	PPB	08 Jun 2023 11:43:15	U	SMPL
PB153279BL - 1	PBW	-239	-0.0671	-	7470A	PPB	08 Jun 2023 11:45:34	U	SMPL
PB153279BS - 1	LCSW	18390	3.9977	-	7470A	PPB	08 Jun 2023 11:50:26	U	SMPL
O3026-02 - 1	VNJ-223	-90	-0.0346	-	7470A	PPB	08 Jun 2023 11:52:42	U	SMPL
O3036-02 - 1	PL-STONE-0602	-53	-0.0266	-	7470A	PPB	08 Jun 2023 11:54:59	U	SMPL
O3039-06 - 1	WC31	-15	-0.0183	-	7470A	PPB	08 Jun 2023 11:57:18	U	SMPL
O3044-04 - 1	SB-02-COMP	53	-0.0034	-	7470A	PPB	08 Jun 2023 11:59:35	U	SMPL
O3044-07 - 1	SB-03-COMP	52	-0.0036	-	7470A	PPB	08 Jun 2023 12:01:50	U	SMPL
CCV02 - 1	CCV02	24164	5.2575	-	7470A	PPB	08 Jun 2023 12:04:06	U	SMPL
CCB02 - 1	CCB02	-237	-0.0667	-	7470A	PPB	08 Jun 2023 12:06:22	U	SMPL
O3044-10 - 1	SB-04-COMP	-51	-0.0261	-	7470A	PPB	08 Jun 2023 12:08:39	U	SMPL
O3044-13 - 1	SB-07-COMP	6	-0.0137	-	7470A	PPB	08 Jun 2023 12:10:54	U	SMPL
O3044-16 - 1	SB-08-COMP	0	-0.0150	-	7470A	PPB	08 Jun 2023 12:13:10	U	SMPL
O3047-03 - 1	WC32	-15	-0.0183	-	7470A	PPB	08 Jun 2023 12:15:26	U	SMPL
O3047-06 - 1	WC33	-13	-0.0178	-	7470A	PPB	08 Jun 2023 12:17:42	U	SMPL
O3050-02 - 1	TP-22-0A	-10	-0.0172	-	7470A	PPB	08 Jun 2023 12:19:59	U	SMPL
O3050-05 - 1	TP-23-0A	0	-0.0150	-	7470A	PPB	08 Jun 2023 12:22:17	U	SMPL
O3069-03 - 1	WC34	16	-0.0115	-	7470A	PPB	08 Jun 2023 12:24:34	U	SMPL
O3069-06 - 1	WC35	19	-0.0108	-	7470A	PPB	08 Jun 2023 12:26:51	U	SMPL
O3069-08 - 1	WC36	4	-0.0141	-	7470A	PPB	08 Jun 2023 12:29:07	U	SMPL
CCV03 - 1	CCV03	24180	5.2610	-	7470A	PPB	08 Jun 2023 12:31:23	U	SMPL
CCB03 - 1	CCB03	-141	-0.0458	-	7470A	PPB	08 Jun 2023 12:33:39	U	SMPL
O3069-08DUP - 1	WC36DUP	-4	-0.0159	-	7470A	PPB	08 Jun 2023 12:35:56	U	SMPL
O3069-08MS - 1	WC36MS	16347	3.5519	-	7470A	PPB	08 Jun 2023 12:38:13	U	SMPL
O3069-08MSD - 1	WC36MSD	16145	3.5078	-	7470A	PPB	08 Jun 2023 12:40:29	U	SMPL
PB153301BL - 1	PBW	100	0.0068	-	7470A	PPB	08 Jun 2023 12:42:46	U	SMPL
PB153301BS - 1	LCSW	16240	3.5285	-	7470A	PPB	08 Jun 2023 12:47:49	U	SMPL
O3066-01 - 1	606	11	-0.0126	-	7470A	PPB	08 Jun 2023 12:50:05	U	SMPL
O3069-09 - 1	TWP08	493	0.0926	-	7470A	PPB	08 Jun 2023 12:57:18	U	SMPL
O3069-09DUP - 1	TWP08DUP	378	0.0675	-	7470A	PPB	08 Jun 2023 12:59:35	U	SMPL
O3069-09MS - 1	TWP08MS	15990	3.4740	-	7470A	PPB	08 Jun 2023 13:01:53	U	SMPL
O3069-09MSD - 1	TWP08MSD	15456	3.3575	-	7470A	PPB	08 Jun 2023 13:04:09	U	SMPL
CCV04 - 1	CCV04	24438	5.3173	-	7470A	PPB	08 Jun 2023 13:06:26	U	SMPL
CCB04 - 1	CCB04	76	0.0016	-	7470A	PPB	08 Jun 2023 13:08:43	U	SMPL
O3069-10 - 1	TWP08	60	-0.0019	-	7470A	PPB	08 Jun 2023 13:11:00	U	SMPL
PB153224TB - 1	PB153224TB	42	-0.0058	-	7470A	PPB	08 Jun 2023 13:13:16	U	SMPL
O3069-08LX5 - 1		-4	-0.0159	-	7470A	PPB	08 Jun 2023 13:15:31	U	SMPL
O3069-08A - 1		21282	4.6287	-	7470A	PPB	08 Jun 2023 13:17:47	U	SMPL
O3069-09LX5 - 1		3	-0.0143	-	7470A	PPB	08 Jun 2023 13:20:03	U	SMPL
O3069-09A - 1		21864	4.7557	-	7470A	PPB	08 Jun 2023 13:22:21	U	SMPL
CCV05 - 1	CCV05	24201	5.2656	-	7470A	PPB	08 Jun 2023 13:24:38	U	SMPL
CCB05 - 1	CCB05	55	-0.0030	-	7470A	PPB	08 Jun 2023 13:26:56	U	SMPL