



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 6/9/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 13:55  
In Date: 06/08/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:50  
Out Date: 06/09/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB125851

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O3057-01 | GBY63           | 1      | 1.18           | 7.46         | 8.64                    | 8.36                      | 96.2    |          |
| O3057-02 | GBY64           | 2      | 1.19           | 7.12         | 8.31                    | 8.25                      | 99.2    |          |
| O3057-03 | GBY65           | 3      | 1.16           | 8.48         | 9.64                    | 9.54                      | 98.8    |          |
| O3057-04 | GBY66           | 4      | 1.14           | 8.57         | 9.71                    | 9.63                      | 99.1    |          |
| O3057-05 | GBY67           | 5      | 1.13           | 6.31         | 7.44                    | 7.32                      | 98.1    |          |
| O3057-06 | GBY68           | 6      | 1.11           | 4.76         | 5.87                    | 5.81                      | 98.7    |          |
| O3057-07 | GBY69           | 7      | 1.19           | 8.44         | 9.63                    | 9.52                      | 98.7    |          |
| O3057-08 | GBY70           | 8      | 1.18           | 7.12         | 8.3                     | 8.24                      | 99.2    |          |
| O3057-09 | GBY71           | 9      | 1.12           | 8.38         | 9.5                     | 9.47                      | 99.6    |          |
| O3057-10 | GBY72           | 10     | 1.14           | 8.49         | 9.63                    | 9.56                      | 99.2    |          |
| O3057-11 | GBY72D          | 11     | 1.14           | 8.49         | 9.63                    | 9.56                      | 99.2    |          |
| O3057-12 | GBY72S          | 12     | 1.14           | 8.49         | 9.63                    | 9.56                      | 99.2    |          |
| O3057-13 | GBY73           | 13     | 1.19           | 8.42         | 9.61                    | 9.54                      | 99.2    |          |
| O3057-14 | GBY74           | 14     | 1.16           | 8.67         | 9.83                    | 9.78                      | 99.4    |          |
| O3057-15 | GBY75           | 15     | 1.18           | 8.59         | 9.77                    | 9.74                      | 99.7    |          |
| O3057-16 | GBY76           | 16     | 1.13           | 8.66         | 9.79                    | 9.7                       | 99.0    |          |
| O3057-17 | GBY77           | 17     | 1.16           | 7.98         | 9.14                    | 9.04                      | 98.7    |          |
| O3057-18 | GBY78           | 18     | 1.19           | 8.62         | 9.81                    | 9.8                       | 99.9    |          |
| O3057-19 | GBY79           | 19     | 1.15           | 8.48         | 9.63                    | 9.55                      | 99.1    |          |
| O3057-20 | GBY80           | 20     | 1.19           | 8.53         | 9.72                    | 9.6                       | 98.6    |          |
| O3057-21 | GBY81           | 21     | 1.17           | 8.55         | 9.72                    | 9.68                      | 99.5    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$