



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 6/9/2023

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:20  
In Date: 06/08/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:00  
Out Date: 06/09/2023  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB125852

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O3058-01 | GBY82           | 1      | 1.16           | 8.50         | 9.66                    | 9.62                      | 99.5    |          |
| O3058-02 | GBY83           | 2      | 1.12           | 8.37         | 9.49                    | 9.37                      | 98.6    |          |
| O3058-03 | GBY84           | 3      | 1.18           | 8.49         | 9.67                    | 9.61                      | 99.3    |          |
| O3058-04 | GBY85           | 4      | 1.12           | 8.65         | 9.77                    | 9.71                      | 99.3    |          |
| O3058-05 | GBY86           | 5      | 1.16           | 8.45         | 9.61                    | 9.56                      | 99.4    |          |
| O3058-06 | GBY87           | 6      | 1.15           | 5.63         | 6.78                    | 6.66                      | 97.9    |          |
| O3058-07 | GBY88           | 7      | 1.18           | 8.50         | 9.68                    | 9.63                      | 99.4    |          |
| O3058-08 | GBY89           | 8      | 1.17           | 8.49         | 9.66                    | 9.63                      | 99.6    |          |
| O3058-09 | GBY90           | 9      | 1.15           | 8.41         | 9.56                    | 9.54                      | 99.8    |          |
| O3058-10 | GBY91           | 10     | 1.19           | 8.55         | 9.74                    | 9.7                       | 99.5    |          |
| O3058-11 | GBY92           | 11     | 1.19           | 8.56         | 9.75                    | 9.7                       | 99.4    |          |
| O3058-12 | GBY92D          | 12     | 1.19           | 8.56         | 9.75                    | 9.7                       | 99.4    |          |
| O3058-13 | GBY92S          | 13     | 1.19           | 8.56         | 9.75                    | 9.7                       | 99.4    |          |
| O3058-14 | GBY93           | 14     | 1.15           | 8.54         | 9.69                    | 9.61                      | 99.1    |          |
| O3058-15 | GBY94           | 15     | 1.15           | 8.38         | 9.53                    | 9.5                       | 99.6    |          |
| O3058-16 | GBY95           | 16     | 1.14           | 8.40         | 9.54                    | 9.5                       | 99.5    |          |
| O3058-17 | GBY96           | 17     | 1.17           | 8.70         | 9.87                    | 9.82                      | 99.4    |          |
| O3058-18 | GBY97           | 18     | 1.13           | 8.80         | 9.93                    | 9.88                      | 99.4    |          |
| O3058-19 | GBY98           | 19     | 1.17           | 8.75         | 9.92                    | 9.89                      | 99.7    |          |
| O3058-20 | GBY99           | 20     | 1.12           | 8.80         | 9.92                    | 9.88                      | 99.5    |          |
| O3058-21 | GBYA0           | 21     | 1.00           | 1.00         | 2.00                    | 1.98                      | 98.0    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$