



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125858

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	06/08/2023	14:20
2	O3039-09		1	12	100.00	>212.0	>212.0	06/08/2023	14:50
3	O3039-09DUP		1	13	100.00	>212.0	>212.0	06/08/2023	15:20
4	O3066-01		1	12	100.00	>212.0	>212.0	06/08/2023	15:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6125858

WorkList Name : fp 6-7

WorkList ID : 170616

Department : Wet-Chemistry

Date : 06-07-2023 08:27:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3039-09	MW-34	Water	Flash Point	Cool 4 deg C	loui01	D11	06/02/2023	1010B
O3066-01	606	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/06/2023	1010B

Date/Time 06/08/2023 14:15
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: HGOON

Date/Time 06/08/2023 16:25
 Raw Sample Received by: HGOON
 Raw Sample Relinquished by: RM CWC