



Analytical Summary Report

Analysis Method: 9041A

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125861

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/08/2023	15:05
2	CAL2	1	Water	NA	NA	20.7	7.00	06/08/2023	15:06
3	CAL3	1	Water	NA	NA	20.7	10.02	06/08/2023	15:07
4	ICV	1	Water	NA	NA	20.8	6.99	06/08/2023	15:10
5	CCV1	1	Water	NA	NA	20.7	2.01	06/08/2023	15:15
6	O3109-01	1	Solid	20.02	20	23.5	8.50	06/08/2023	15:30
7	CCV2	1	Water	NA	NA	20.8	12.02	06/08/2023	15:35

UP 125861

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S 03109 WorkList ID : 170699 Department : Wet-Chemistry Date : 06-08-2023 14:27:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3109-01	SP-1-COMP	Solid	pH	Cool 4 deg C	EART12	G11	06/08/2023	9041A

Date/Time 06/09/23 15:00
Raw Sample Received by: Ad Sin
Raw Sample Relinquished by: Ad Sin

Date/Time 06/09/23 17:00
Raw Sample Received by: CP Sin
Raw Sample Relinquished by: Ad Sin