

LB125863


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41955



A= 0.0000e+000

B= 2.3176e-004 slope

C= -7.6273e-002 $y = mx + c$

Rho= 0.9981786

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.056	-0.056	89	0.000	89					-10
0.05	0.050					302	0				-5
0.20	0.200	0.180	-0.020	1105	0.0 %	1105					+5
2.50	2.500	2.376	-0.124	10582	0.0 %	10582					+4
5.00	5.000	5.240	0.240	22941	0.0 %	22941					-4
7.50	7.500	7.811	0.311	34033	0.0 %	34033					
10.0	10.000	9.648	-0.352	41959	0.0 %	41959					

LB125863

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	89	-		0 SFAM01.1	PPB	6/8/2023 15:01	Std
	0.2 S01	1105	-		0.2 SFAM01.1	PPB	6/8/2023 15:03	Std
	2.5 S02	10582	-		2.5 SFAM01.1	PPB	6/8/2023 15:06	Std
	5 S03	22941	-		5 SFAM01.1	PPB	6/8/2023 15:16	Std
	7.5 S04	34033	-		7.5 SFAM01.1	PPB	6/8/2023 15:21	Std
	10 S05	41959	-		10 SFAM01.1	PPB	6/8/2023 15:28	Std
ICV013	ICV013	16237	3.6868 -		SFAM01.1	PPB	6/8/2023 15:36	SMPL
ICB013	ICB013	-18	-0.0804 -		SFAM01.1	PPB	6/8/2023 15:38	SMPL
CCV006	CCV006	22068	5.0382 -		SFAM01.1	PPB	6/8/2023 15:40	SMPL
CCB006	CCB006	-41	-0.0858 -		SFAM01.1	PPB	6/8/2023 15:42	SMPL
PB153308BL	PBS308	24	-0.0707 -		SFAM01.1	PPB	6/8/2023 15:45	SMPL
O3052-01	GBXD4	2693	0.5478 -		SFAM01.1	PPB	6/8/2023 15:47	SMPL
O3052-02	GBXD5	334	0.0011 -		SFAM01.1	PPB	6/8/2023 15:49	SMPL
O3052-03	GBXD6	517	0.0435 -		SFAM01.1	PPB	6/8/2023 15:51	SMPL
O3052-04	GBXD7	511	0.0422 -		SFAM01.1	PPB	6/8/2023 15:54	SMPL
O3052-05	GBXD8	181	-0.0343 -		SFAM01.1	PPB	6/8/2023 15:56	SMPL
O3052-06	GBXD9	152	-0.041 -		SFAM01.1	PPB	6/8/2023 15:58	SMPL
O3052-07	GBXE0	387	0.0134 -		SFAM01.1	PPB	6/8/2023 16:01	SMPL
O3052-08	GBXE1	473	0.0333 -		SFAM01.1	PPB	6/8/2023 16:03	SMPL
O3052-09	GBXE2	162	-0.0387 -		SFAM01.1	PPB	6/8/2023 16:05	SMPL
O3052-11	GBXE4	547	0.0505 -		SFAM01.1	PPB	6/8/2023 16:10	SMPL
O3052-12	GBXE5	196	-0.0308 -		SFAM01.1	PPB	6/8/2023 16:12	SMPL
O3052-13	GBXE5D	221	-0.0251 -		SFAM01.1	PPB	6/8/2023 16:14	SMPL
O3052-10	GBXE3	192	-0.0318 -		SFAM01.1	PPB	6/8/2023 16:16	SMPL
O3052-15	GBXE6	517	0.0435 -		SFAM01.1	PPB	6/8/2023 16:21	SMPL
O3052-16	GBXE7	2688	0.5467 -		SFAM01.1	PPB	6/8/2023 16:24	SMPL
O3052-17	GBXE8	841	0.1186 -		SFAM01.1	PPB	6/8/2023 16:26	SMPL
O3052-18	GBXE9	3449	0.7231 -		SFAM01.1	PPB	6/8/2023 16:28	SMPL
O3052-19	GBXF0	4451	0.9553 -		SFAM01.1	PPB	6/8/2023 16:31	SMPL
CCV007	CCV007	22907	5.2326 -		SFAM01.1	PPB	6/8/2023 16:33	SMPL
CCB007	CCB007	-28	-0.0828 -		SFAM01.1	PPB	6/8/2023 16:35	SMPL
O3052-20	GBXF1	475	0.0338 -		SFAM01.1	PPB	6/8/2023 16:37	SMPL
O3052-21	GBXF2	723	0.0913 -		SFAM01.1	PPB	6/8/2023 16:40	SMPL
PB153309BL	PBS309	1	-0.076 -		SFAM01.1	PPB	6/8/2023 16:42	SMPL
O3053-01	GBXF3	540	0.0489 -		SFAM01.1	PPB	6/8/2023 16:44	SMPL
O3053-02	GBXF4	17506	3.9809 -		SFAM01.1	PPB	6/8/2023 16:46	SMPL
O3053-03	GBXF5	13614	3.0789 -		SFAM01.1	PPB	6/8/2023 16:49	SMPL
O3053-04	GBXF6	18142	4.1283 -		SFAM01.1	PPB	6/8/2023 16:51	SMPL
O3053-05	GBXF7	168	-0.0373 -		SFAM01.1	PPB	6/8/2023 16:53	SMPL
O3053-06	GBXF8	240	-0.0207 -		SFAM01.1	PPB	6/8/2023 16:56	SMPL
O3053-07	GBXF9	114	-0.0499 -		SFAM01.1	PPB	6/8/2023 16:58	SMPL
O3053-08	GBXF9D	21	-0.0714 -		SFAM01.1	PPB	6/8/2023 17:00	SMPL
O3053-09	GBXF9S	12464	2.8124 -		SFAM01.1	PPB	6/8/2023 17:02	SMPL
O3053-10	GBXG0	16	-0.0726 -		SFAM01.1	PPB	6/8/2023 17:05	SMPL
O3053-11	GBXG1	27	-0.07 -		SFAM01.1	PPB	6/8/2023 17:07	SMPL
O3053-12	GBXG2	7	-0.0747 -		SFAM01.1	PPB	6/8/2023 17:09	SMPL