



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125869

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/08/2023	15:05
2	CAL2	1	Water	NA	NA	20.7	7.00	06/08/2023	15:06
3	CAL3	1	Water	NA	NA	20.7	10.02	06/08/2023	15:07
4	ICV	1	Water	NA	NA	20.8	6.99	06/08/2023	15:10
5	CCV1	1	Water	NA	NA	20.7	2.01	06/08/2023	15:15
6	O3096-02	1	Solid	20.02	20	21.0	8.36	06/08/2023	15:20
7	O3096-05	1	Solid	20.03	20	20.6	8.76	06/08/2023	15:25
8	O3109-02	1	Solid	20.02	20	23.5	8.50	06/08/2023	15:30
9	O3109-02DUP	1	Solid	20.01	20	23.6	8.51	06/08/2023	15:31
10	CCV2	1	Water	NA	NA	20.8	12.02	06/08/2023	15:35

125869

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRRSIVITY O3096

WorkList ID : 170688

Department : Wet-Chemistry

Date : 06-08-2023 11:41:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3096-02	TP-24-0A	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/07/2023	9045D
O3096-05	TP-25-0A	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/07/2023	9045D
O3109-02	SP-1-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	G11	06/08/2023	9045D

Date/Time 06/08/23 15:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/08/23 17:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]