

LB125873


7471B

INSTRUMENT ID: C41

Linear

 μ Abs.:

56153



A= 0.0000e+000

B= 1.7529e-004

C= -7.4136e-002

Rho= 0.9991641

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	401
0.0	0.000	-0.036	-0.036	220	0.000	220					1
0.2	0.200	0.189	-0.011	1501	0.0 %	1501					-6
2.5	2.500	2.403	-0.097	14131	0.0 %	14131					-4
5.0	5.000	5.157	0.157	29841	0.0 %	29841					-3
7.5	7.500	7.718	0.218	44456	0.0 %	44456					-3
10.0	10.000	9.769	-0.231	56153	0.0 %	56153					-2

LB125873 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Jun 2023 10:36:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	220	-	0.0000	7471B	PPB	09 Jun 2023 11:01:03	S	Std
0.2 - 1	50.2	1501	-	0.2000	7471B	PPB	09 Jun 2023 11:03:21	S	Std
2.5 - 1	52.5	14131	-	2.5000	7471B	PPB	09 Jun 2023 11:08:16	S	Std
5.0 - 1	55	29841	-	5.0000	7471B	PPB	09 Jun 2023 11:14:44	S	Std
7.5 - 1	57.5	44456	-	7.5000	7471B	PPB	09 Jun 2023 11:19:36	S	Std
10.0 - 1	516	56153	-	10.0000	7471B	PPB	09 Jun 2023 11:24:35	S	Std
ICV14 - 1	ICV14	24557	4.2304	-	7471B	PPB	09 Jun 2023 11:27:29	U	SMPL
ICB14 - 1	ICB14	65	-0.0627	-	7471B	PPB	09 Jun 2023 11:29:44	U	SMPL
CCV11 - 1	CCV11	30658	5.2998	-	7471B	PPB	09 Jun 2023 11:32:01	U	SMPL
CCB11 - 1	CCB11	58	-0.0640	-	7471B	PPB	09 Jun 2023 11:34:16	U	SMPL
CRA - 1	CRA	1521	0.1925	-	7471B	PPB	09 Jun 2023 11:36:34	U	SMPL
HighStd - 1	HighStd	56478	9.8257	-	7471B	PPB	09 Jun 2023 11:38:49	U	SMPL
ChkStd - 1	ChkStd	41478	7.1964	-	7471B	PPB	09 Jun 2023 11:41:05	U	SMPL
PB153313BL - 1	PBS	-53	-0.0834	-	7471B	PPB	09 Jun 2023 11:43:24	U	SMPL
PB153313BS - 1	LCSS	24047	4.1410	-	7471B	PPB	09 Jun 2023 11:45:43	U	SMPL
O3069-01 - 1	SB34A	18520	3.1722	-	7471B	PPB	09 Jun 2023 11:52:59	U	SMPL
O3069-01DUP - 1	SB34ADUP	18040	3.0880	-	7471B	PPB	09 Jun 2023 11:55:15	U	SMPL
O3069-01MS - 1	SB34AMS	38498	6.6741	-	7471B	PPB	09 Jun 2023 11:57:35	U	SMPL
O3069-01MSD - 1	SB34AMSD	40033	6.9431	-	7471B	PPB	09 Jun 2023 12:02:20	U	SMPL
O3069-02 - 1	SB34B	710	0.0503	-	7471B	PPB	09 Jun 2023 12:04:37	U	SMPL
CCV12 - 1	CCV12	29506	5.0979	-	7471B	PPB	09 Jun 2023 12:06:55	U	SMPL
CCB12 - 1	CCB12	117	-0.0536	-	7471B	PPB	09 Jun 2023 12:09:10	U	SMPL
O3069-04 - 1	SB35A	15090	2.5709	-	7471B	PPB	09 Jun 2023 12:11:28	U	SMPL
O3069-05 - 1	SB35B	22127	3.8044	-	7471B	PPB	09 Jun 2023 12:13:43	U	SMPL
O3069-07 - 1	SB36	3279	0.5006	-	7471B	PPB	09 Jun 2023 12:16:01	U	SMPL
O3071-01 - 1	WC-1	5649	0.9161	-	7471B	PPB	09 Jun 2023 12:18:19	U	SMPL
O3072-01 - 1	30999	1749	0.2324	-	7471B	PPB	09 Jun 2023 12:20:35	U	SMPL
O3072-02 - 1	30100	343	-0.0140	-	7471B	PPB	09 Jun 2023 12:22:52	U	SMPL
O3072-03 - 1	30101	5725	0.9294	-	7471B	PPB	09 Jun 2023 12:25:08	U	SMPL
O3086-01 - 1	HD-1-060723	2912	0.4363	-	7471B	PPB	09 Jun 2023 12:27:25	U	SMPL
O3092-01 - 1	TP-1	639	0.0379	-	7471B	PPB	09 Jun 2023 12:29:43	U	SMPL
O3092-04 - 1	TP-2	769	0.0607	-	7471B	PPB	09 Jun 2023 12:31:58	U	SMPL
CCV13 - 1	CCV13	30700	5.3072	-	7471B	PPB	09 Jun 2023 12:34:14	U	SMPL
CCB13 - 1	CCB13	127	-0.0519	-	7471B	PPB	09 Jun 2023 12:36:30	U	SMPL
O3093-01 - 1	SU-3-06072023	3258	0.4969	-	7471B	PPB	09 Jun 2023 12:38:48	U	SMPL
O3094-01 - 1	WC-1	9134	1.5269	-	7471B	PPB	09 Jun 2023 12:41:03	U	SMPL
O3096-01 - 1	TP-24-0A	682	0.0454	-	7471B	PPB	09 Jun 2023 12:43:19	U	SMPL
O3096-04 - 1	TP-25-0A	772	0.0612	-	7471B	PPB	09 Jun 2023 12:45:36	U	SMPL
O3103-01 - 1	NB-8-060823	753	0.0579	-	7471B	PPB	09 Jun 2023 12:47:52	U	SMPL
O3109-01 - 1	SP-1-COMP	55321	9.6229	-	7471B	PPB	09 Jun 2023 12:50:08	U	SMPL
O3069-01LX5 - 1		3328	0.5092	-	7471B	PPB	09 Jun 2023 12:52:25	U	SMPL
O3069-01A - 1		41734	7.2413	-	7471B	PPB	09 Jun 2023 12:54:45	U	SMPL
CCV14 - 1	CCV14	30326	5.2416	-	7471B	PPB	09 Jun 2023 12:59:20	U	SMPL
CCB14 - 1	CCB14	79	-0.0603	-	7471B	PPB	09 Jun 2023 13:01:36	U	SMPL