



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB125878
Thermometer ID: Flashpoint

Reviewed By: Rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 755.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	06/09/2023	12:20
2	O3046-01		1	12	100.00	>212.0	>212.0	06/09/2023	12:50
3	O3046-01DUP		1	13	100.00	>212.0	>212.0	06/09/2023	13:20
4	O3046-07		1	12	100.00	>212.0	>212.0	06/09/2023	13:50
5	O3046-13		1	14	100.00	>212.0	>212.0	06/09/2023	14:20
6	O3047-09		1	12	100.00	>212.0	>212.0	06/09/2023	14:50
7	O3047-12		1	12	100.00	>212.0	>212.0	06/09/2023	15:20
8	O3069-11		1	13	100.00	>212.0	>212.0	06/09/2023	15:50
9	O3072-04		1	14	100.00	>212.0	>212.0	06/09/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6125878

WorkList Name : fp 6-9

WorkList ID : 170719

Department : Wet-Chemistry

Date : 06-09-2023 08:23:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3046-01	COMP-1-1-5	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/05/2023	1010B
O3046-07	COMP-2-6-10	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/05/2023	1010B
O3046-13	COMP-3-11-15	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/05/2023	1010B
O3047-09	MW-41	Water	Flash Point	Cool 4 deg C	loui01	D11	06/05/2023	1010B
O3047-12	TWP-07	Water	Flash Point	Cool 4 deg C	loui01	D11	06/05/2023	1010B
O3069-11	TWP08	Water	Flash Point	Cool 4 deg C	loui01	D11	06/06/2023	1010B
O3072-04	30998	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/07/2023	1010B

Date/Time 06/16/2023 12:00
 Raw Sample Received by: RM CWS
 Raw Sample Relinquished by: JWC

Date/Time 06/16/2023 16:55
 Raw Sample Received by: JWC
 Raw Sample Relinquished by: JWC