



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/12/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/09/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 06/10/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125879

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3114-01	MARION-SOIL-0609	19	1.18	8.45	9.63	9.04	93.0	
O3119-05	SVOC-GPC-BLANK	1	1.00	1.00	2.00	2.00	100.0	
O3119-06	PEST-GPC-BLANK	2	1.00	1.00	2.00	2.00	100.0	
O3119-07	PEST-GPC-BLANK-SPIKE	3	1.00	1.00	2.00	2.00	100.0	
O3119-08	PCB-GPC-BLANK	4	1.00	1.00	2.00	2.00	100.0	
O3119-09	PCB-GPC-BLANK-SPIKE	5	1.00	1.00	2.00	2.00	100.0	
O3119-10	SVOC-GPC2-BLANK	6	1.00	1.00	2.00	2.00	100.0	
O3119-11	PEST-GPC2-BLANK	7	1.00	1.00	2.00	2.00	100.0	
O3119-12	PEST-GPC2-BLANK-SPIKE	8	1.00	1.00	2.00	2.00	100.0	
O3119-13	PCB-GCP2-BLANK	9	1.00	1.00	2.00	2.00	100.0	
O3119-14	PCB-GCP2-BLANK-SPIKE	10	1.00	1.00	2.00	2.00	100.0	
O3120-01	SS055P-MW02S-SB01-00G	11	1.14	8.54	9.68	9.12	93.4	
O3120-02	SS055P-MW02D-SB01-00G	12	1.19	8.48	9.67	9.02	92.3	
O3120-03	SS055P-MW04-SB01-00G	13	1.11	8.78	9.89	8.57	85.0	
O3123-15	BP-VPB-RW11-GW-738-740	14	1.15	8.39	9.54	1.75	7.2	SLUDGE SAMPLE
O3123-16	BP-VPB-RW11-GW-838-840	15	1.15	8.84	9.99	2.08	10.5	SLUDGE SAMPLE
O3123-17	BP-VPB-RW11-GW-823-825	16	1.18	8.79	9.97	2.14	10.9	SLUDGE SAMPLE
O3138-01	WC-1	17	1.15	8.80	9.95	9.11	90.5	
O3138-02	WC-1-EPH	18	1.12	8.60	9.72	8.92	90.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$