


43125882

SFAM01.1

INSTRUMENT 101 CV1

μ Abs.:

56371



Linear

A= 0.0000e+000

B= 1.7542e-004 slope

C= 1.0719e-002 intercept

Rho= 0.9995419

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	D/D1)
0.00	0.000	0.020	0.020	55	0.000	55					
0.05	0.050					302	0				
0.20	0.200	0.193	-0.007	1040	0.0 %	1040					-5 -4 NB
2.50	2.500	2.346	-0.154	13312	0.0 %	13312					-6 6/16/12
5.00	5.000	5.196	0.196	29561	0.0 %	29561					4
7.50	7.500	7.544	0.044	42946	0.0 %	42946					1
10.0	10.000	9.900	-0.100	56371	0.0 %	56371					-1

LB125882

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	55	-		0 SFAM01.1	PPB	6/9/2023 14:30	Std
	0.2 S01	1040	-		0.2 SFAM01.1	PPB	6/9/2023 14:32	Std
	2.5 S02	13312	-		2.5 SFAM01.1	PPB	6/9/2023 14:34	Std
	5 S03	29561	-		5 SFAM01.1	PPB	6/9/2023 14:37	Std
	7.5 S04	42946	-		7.5 SFAM01.1	PPB	6/9/2023 14:39	Std
	10 S05	56371	-		10 SFAM01.1	PPB	6/9/2023 14:44	Std
ICV015	ICV015	22329	3.9278 -		SFAM01.1	PPB	6/9/2023 14:47	SMPL
ICB015	ICB015	-24	0.0065 -		SFAM01.1	PPB	6/9/2023 14:49	SMPL
CCV015	CCV015	29061	5.1087 -		SFAM01.1	PPB	6/9/2023 14:51	SMPL
CCB015	CCB015	-18	0.0076 -		SFAM01.1	PPB	6/9/2023 14:54	SMPL
PB153341BL	PBS341	30	0.016 -		SFAM01.1	PPB	6/9/2023 14:56	SMPL
O3055-01	GBXK1	975	0.1818 -		SFAM01.1	PPB	6/9/2023 14:58	SMPL
O3055-02	GBXK2	398	0.0805 -		SFAM01.1	PPB	6/9/2023 15:00	SMPL
O3055-03	GBXK3	578	0.1121 -		SFAM01.1	PPB	6/9/2023 15:03	SMPL
O3055-04	GBXK4	2109	0.3807 -		SFAM01.1	PPB	6/9/2023 15:05	SMPL
O3055-05	GBXK5	346	0.0714 -		SFAM01.1	PPB	6/9/2023 15:07	SMPL
O3055-06	GBXK6	787	0.1488 -		SFAM01.1	PPB	6/9/2023 15:09	SMPL
O3055-07	GBXK7	1299	0.2386 -		SFAM01.1	PPB	6/9/2023 15:12	SMPL
O3055-08	GBXK8	1525	0.2782 -		SFAM01.1	PPB	6/9/2023 15:14	SMPL
O3055-09	GBXK9	280	0.0598 -		SFAM01.1	PPB	6/9/2023 15:16	SMPL
O3055-10	GBXL0	387	0.0786 -		SFAM01.1	PPB	6/9/2023 15:19	SMPL
O3055-11	GBXL0D	374	0.0763 -		SFAM01.1	PPB	6/9/2023 15:24	SMPL
O3055-12	GBXL0S	15050	2.6509 -		SFAM01.1	PPB	6/9/2023 15:26	SMPL
O3055-13	GBXL1	735	0.1397 -		SFAM01.1	PPB	6/9/2023 15:28	SMPL
O3055-14	GBXL2	191	0.0442 -		SFAM01.1	PPB	6/9/2023 15:30	SMPL
O3055-16	GBXL4	20608	3.6259 -		SFAM01.1	PPB	6/9/2023 15:33	SMPL
O3055-17	GBXL5	12127	2.1381 -		SFAM01.1	PPB	6/9/2023 15:35	SMPL
O3055-18	GBXL6	13598	2.3961 -		SFAM01.1	PPB	6/9/2023 15:37	SMPL
O3055-19	GBXL7	1091	0.2021 -		SFAM01.1	PPB	6/9/2023 15:40	SMPL
O3055-20	GBXL8	370	0.0756 -		SFAM01.1	PPB	6/9/2023 15:42	SMPL
CCV016	CCV016	28701	5.0456 -		SFAM01.1	PPB	6/9/2023 15:44	SMPL
CCB016	CCB016	-36	0.0044 -		SFAM01.1	PPB	6/9/2023 15:46	SMPL
O3055-21	GBXL9	410	0.0826 -		SFAM01.1	PPB	6/9/2023 15:49	SMPL
PB153342BL	PBS153342	12	0.0128 -		SFAM01.1	PPB	6/9/2023 15:51	SMPL
O3056-01	GBXM0	329	0.0684 -		SFAM01.1	PPB	6/9/2023 15:53	SMPL
O3056-02	GBXM1	9324	1.6464 -		SFAM01.1	PPB	6/9/2023 15:55	SMPL
O3056-03	GBXM2	4138	0.7366 -		SFAM01.1	PPB	6/9/2023 15:58	SMPL
O3056-04	GBXM3	2225	0.401 -		SFAM01.1	PPB	6/9/2023 16:00	SMPL
O3056-05	GBXM4	3670	0.6545 -		SFAM01.1	PPB	6/9/2023 16:02	SMPL
O3056-06	GBXM5	1993	0.3603 -		SFAM01.1	PPB	6/9/2023 16:05	SMPL
O3056-07	GBXM5D	1810	0.3282 -		SFAM01.1	PPB	6/9/2023 16:07	SMPL
O3056-08	GBXM5S	15104	2.6603 -		SFAM01.1	PPB	6/9/2023 16:09	SMPL
O3056-09	GBXM6	27935	4.9112 -		SFAM01.1	PPB	6/9/2023 16:16	SMPL
CCV017	CCV017	28507	5.0115 -		SFAM01.1	PPB	6/9/2023 16:21	SMPL
CCB017	CCB017	-89	-0.0049 -		SFAM01.1	PPB	6/9/2023 16:23	SMPL