



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125884

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/09/2023	15:15
2	CAL2	1	Water	NA	NA	20.8	6.99	06/09/2023	15:16
3	CAL3	1	Water	NA	NA	20.7	10.02	06/09/2023	15:19
4	ICV	1	Water	NA	NA	20.8	7.00	06/09/2023	15:20
5	CCV1	1	Water	NA	NA	20.8	2.01	06/09/2023	15:22
6	O3116-01	1	Water	NA	NA	23.1	7.55	06/09/2023	15:25
7	O3116-01DUP	1	Water	NA	NA	23.2	7.56	06/09/2023	15:26
8	O3137-01	1	Water	NA	NA	25.6	11.21	06/09/2023	15:30
9	CCV2	1	Water	NA	NA	20.7	12.02	06/09/2023	15:33

✓ 125884

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o3137

WorkList ID : 170743

Department : Wet-Chemistry

Date : 06-09-2023 15:01:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3116-01	NWB-1878	Water	pH	Cool 4 deg C	PSEG03	D12	06/09/2023	9040C
O3137-01	ACEDEMY-STREET-DRUM	Water	pH	Cool 4 deg C	PSEG03	D21	06/09/2023	9040C

16:10

Date/Time

06/09/23

Raw Sample Received by:

JP/WCC

Raw Sample Relinquished by:

JP/WCC

16:30

Date/Time

06/09/23

Raw Sample Received by:

JP/WCC

Raw Sample Relinquished by:

JP/WCC